

Surrogate Summary

SDG No.: Q2001

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2001-01	WC-A4-03-G	1,2-Dichloroethane-d4	50	55.0	110	70 (74)	130 (125)
		Dibromofluoromethane	50	42.2	84	70 (75)	130 (124)
		Toluene-d8	50	52.0	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.0	104	70 (77)	130 (121)
Q2001-05	WC-A1-05-G	1,2-Dichloroethane-d4	50	51.6	103	70 (74)	130 (125)
		Dibromofluoromethane	50	40.8	82	70 (75)	130 (124)
		Toluene-d8	50	51.6	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.1	108	70 (77)	130 (121)
Q2001-05DL	WC-A1-05-GDL	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	50.5	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.6	107	70 (77)	130 (121)
Q2001-09	WC-A1-06-G	1,2-Dichloroethane-d4	50	51.3	103	70 (74)	130 (125)
		Dibromofluoromethane	50	42.3	85	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.5	105	70 (77)	130 (121)
Q2001-09DL	WC-A1-06-GDL	1,2-Dichloroethane-d4	50	52.9	106	70 (74)	130 (125)
		Dibromofluoromethane	50	50.2	100	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.6	103	70 (77)	130 (121)
Q2001-13	WC-A1-07-G	1,2-Dichloroethane-d4	50	51.0	102	70 (74)	130 (125)
		Dibromofluoromethane	50	36.4	73	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.4	107	70 (77)	130 (121)
Q2001-13DL	WC-A1-07-GDL	1,2-Dichloroethane-d4	50	52.5	105	70 (74)	130 (125)
		Dibromofluoromethane	50	49.6	99	70 (75)	130 (124)
		Toluene-d8	50	50.9	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106	70 (77)	130 (121)

Surrogate Summary

SDG No.: Q2001

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0513WBL01	VX0513WBL01	1,2-Dichloroethane-d4	50	53.8	108	70 (74)	130 (125)
		Dibromofluoromethane	50	51.5	103	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.2	100	70 (77)	130 (121)
VX0513WBS01	VX0513WBS01	1,2-Dichloroethane-d4	50	51.5	103	70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.3	99	70 (77)	130 (121)
VX0513WBSD0	VX0513WBSD01	1,2-Dichloroethane-d4	50	52.2	104	70 (74)	130 (125)
		Dibromofluoromethane	50	51.9	104	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)
VX0514WBL01	VX0514WBL01	1,2-Dichloroethane-d4	50	53.7	107	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.1	102	70 (77)	130 (121)
VX0514WBS01	VX0514WBS01	1,2-Dichloroethane-d4	50	52.1	104	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.2	100	70 (77)	130 (121)

