

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001
 Lab File ID: VX046153.D Date Analyzed: 05/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98502	5.54	167387	6.75	146749	10.05
UPPER LIMIT	197004	6.043	334774	7.25	293498	10.549
LOWER LIMIT	49251	5.043	83693.5	6.25	73374.5	9.549
EPA SAMPLE NO.						
WC-A4-03-G	60233	5.55	120477	6.76	113675	10.05
WC-A1-05-G	61818	5.55	120158	6.76	115770	10.06
WC-A1-06-G	69385	5.55	132178	6.76	122795	10.06
WC-A1-07-G	63010	5.55	118470	6.76	115809	10.06
VX0513WBL01	67277	5.55	133345	6.76	124650	10.05
VX0513WBS01	89387	5.54	160585	6.76	139963	10.06
VX0513WBSD01	85900	5.54	148923	6.76	132995	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.