

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001
 Lab File ID: VX046153.D Date Analyzed: 05/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70437	12.018				
UPPER LIMIT	140874	12.518				
LOWER LIMIT	35218.5	11.518				
EPA SAMPLE NO.						
WC-A4-03-G	50921	12.02				
WC-A1-05-G	55282	12.02				
WC-A1-06-G	59221	12.02				
WC-A1-07-G	55553	12.02				
VX0513WBL01	52712	12.02				
VX0513WBS01	64268	12.02				
VX0513WBSD01	62827	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.