

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001
Lab File ID: VX046180.D BFB Injection Date: 05/14/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:29
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	57.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.2 (0.3) 1
174	50.0 - 100.0% of mass 95	69
175	5.0 - 9.0% of mass 174	5.2 (7.5) 1
176	95.0 - 101.0% of mass 174	65.9 (95.5) 1
177	5.0 - 9.0% of mass 176	4.1 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046181.D	05/14/2025	09:59
VX0514WBL01	VX0514WBL01	VX046183.D	05/14/2025	10:50
VX0514WBS01	VX0514WBS01	VX046188.D	05/14/2025	12:46
WC-A1-05-GDL	Q2001-05DL	VX046190.D	05/14/2025	13:35
WC-A1-06-GDL	Q2001-09DL	VX046191.D	05/14/2025	13:59
WC-A1-07-GDL	Q2001-13DL	VX046192.D	05/14/2025	14:22