

SOLID EPH_F2 LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Sciacca General Contractors, LLC
Lab Code: CHEM **Cas No:** Q2003 **SAS No :** Q2003 **SDG No:** Q2003
Sample No : PB168029BSD **Datafile:** FG015860.D
Client ID : PB168029BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C28-C40	30.0	13.1	44		4.5 (40-140)	25
Aliphatic C9-C28	99.9	56.2	57		4.1 (40-140)	25