

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: SCIA01
 ProjectID: 32 Sayer Pl, Hillsdale
 Lab Code: CHEM Case No.: Q2003 SAS No.: Q2003 SDG No.: Q2003

Calibration Sequence : FF042225		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	171154261	100679	FF015786.D
850	88482232	104097	FF015787.D
340	33178146	97583	FF015788.D
170	17736847	104334	FF015789.D
85	10850659	127655	FF015790.D
AVG RF : 106870		% RSD : 11.177	AVG RT : 15.02