

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: SCIA01
ProjectID: 32 Sayer Pl, Hillsdale
Lab Code: CHEM Case No.: Q2003 SAS No.: Q2003 SDG No.: Q2003
DataFile: FF015851.D Analyst Name: YP\AJ Analyst Date: 05-16-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	91013278	107074	106870	0.191