

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCIA01

Lab Code: CHEM Case No.: Q2003 SAS No.: Q2003 SDG No.: Q2003

Instrument ID: MSVOA_Y Calibration Date/Time: 05/16/2025 08:24

Lab File ID: VY022277.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.522	0.521		-0.19	20
Chloromethane	1.097	1.067	0.1	-2.73	20
Vinyl Chloride	1.226	1.343		9.54	20
Bromomethane	0.998	1.071		7.32	20
Chloroethane	0.763	0.887		16.25	20
Trichlorofluoromethane	1.241	1.308		5.4	20
1,1,2-Trichlorotrifluoroethane	0.539	0.569		5.57	20
1,1-Dichloroethene	0.535	0.570		6.54	20
Acetone	0.102	0.103		0.98	20
Carbon Disulfide	1.744	1.850		6.08	20
Methyl tert-butyl Ether	1.475	1.543		4.61	20
Methyl Acetate	0.323	0.324		0.31	20
Methylene Chloride	0.662	0.631		-4.68	20
trans-1,2-Dichloroethene	0.590	0.626		6.1	20
1,1-Dichloroethane	1.106	1.172	0.1	5.97	20
Cyclohexane	1.105	1.093		-1.09	20
2-Butanone	0.171	0.166		-2.92	20
Carbon Tetrachloride	0.509	0.537		5.5	20
cis-1,2-Dichloroethene	0.681	0.737		8.22	20
Bromochloromethane	0.472	0.487		3.18	20
Chloroform	1.061	1.161		9.43	20
1,1,1-Trichloroethane	0.974	1.038		6.57	20
Methylcyclohexane	0.660	0.680		3.03	20
Benzene	1.457	1.567		7.55	20
1,2-Dichloroethane	0.394	0.417		5.84	20
Trichloroethene	0.362	0.381		5.25	20
1,2-Dichloropropane	0.351	0.371		5.7	20
Bromodichloromethane	0.492	0.533		8.33	20
4-Methyl-2-Pentanone	0.253	0.250		-1.19	20
Toluene	0.923	0.982		6.39	20
t-1,3-Dichloropropene	0.480	0.500		4.17	20
cis-1,3-Dichloropropene	0.550	0.587		6.73	20
1,1,2-Trichloroethane	0.243	0.254		4.53	20
2-Hexanone	0.170	0.165		-2.94	20
Dibromochloromethane	0.315	0.338		7.3	20
1,2-Dibromoethane	0.226	0.238		5.31	20
Tetrachloroethene	0.462	0.477		3.25	20
Chlorobenzene	1.128	1.200	0.3	6.38	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCIA01

Lab Code: CHEM Case No.: Q2003 SAS No.: Q2003 SDG No.: Q2003

Instrument ID: MSVOA_Y Calibration Date/Time: 05/16/2025 08:24

Lab File ID: VY022277.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	2.113	2.235		5.77	20
m/p-Xylenes	0.796	0.848		6.53	20
o-Xylene	0.753	0.804		6.77	20
Styrene	1.270	1.355		6.69	20
Bromoform	0.209	0.214	0.1	2.39	20
Isopropylbenzene	4.142	4.419		6.69	20
1,1,2,2-Tetrachloroethane	0.637	0.651	0.3	2.2	20
1,3-Dichlorobenzene	1.806	1.921		6.37	20
1,4-Dichlorobenzene	1.737	1.826		5.12	20
1,2-Dichlorobenzene	1.517	1.596		5.21	20
1,2-Dibromo-3-Chloropropane	0.103	0.104		0.97	20
1,2,4-Trichlorobenzene	0.876	0.929		6.05	20
1,2,3-Trichlorobenzene	0.743	0.768		3.37	20
1,2-Dichloroethane-d4	0.546	0.552		1.1	20
Dibromofluoromethane	0.298	0.308		3.36	20
Toluene-d8	1.222	1.265		3.52	20
4-Bromofluorobenzene	0.387	0.388		0.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.