

Surrogate Summary

SDG No.: **Q2003**

Client: **Sciacca General Contractors, LLC**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2003-02	VOC	1,2-Dichloroethane-d4	50	39.4	79	70 (63)	130 (155)
		Dibromofluoromethane	50	44.9	90	70 (70)	130 (134)
		Toluene-d8	50	44.1	88	70 (74)	130 (123)
		4-Bromofluorobenzene	50	28.7	57 *	70 (38)	130 (136)
Q2003-02RE	VOCRE	1,2-Dichloroethane-d4	50	34.9	70	70 (63)	130 (155)
		Dibromofluoromethane	50	46.0	92	70 (70)	130 (134)
		Toluene-d8	50	45.6	91	70 (74)	130 (123)
		4-Bromofluorobenzene	50	29.4	59 *	70 (38)	130 (136)
VY0516SBL01	VY0516SBL01	1,2-Dichloroethane-d4	50	43.9	88	70 (63)	130 (155)
		Dibromofluoromethane	50	48.0	96	70 (70)	130 (134)
		Toluene-d8	50	48.8	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.5	97	70 (38)	130 (136)
VY0516SBS01	VY0516SBS01	1,2-Dichloroethane-d4	50	51.1	102	70 (63)	130 (155)
		Dibromofluoromethane	50	50.8	102	70 (70)	130 (134)
		Toluene-d8	50	50.9	102	70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.4	99	70 (38)	130 (136)
VY0516SBSD01	VY0516SBSD01	1,2-Dichloroethane-d4	50	48.8	98	70 (63)	130 (155)
		Dibromofluoromethane	50	50.0	100	70 (70)	130 (134)
		Toluene-d8	50	50.6	101	70 (74)	130 (123)
		4-Bromofluorobenzene	50	47.3	95	70 (38)	130 (136)
VY0519SBL01	VY0519SBL01	1,2-Dichloroethane-d4	50	41.1	82	70 (63)	130 (155)
		Dibromofluoromethane	50	47.5	95	70 (70)	130 (134)
		Toluene-d8	50	48.8	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	37.7	75	70 (38)	130 (136)
VY0519SBS01	VY0519SBS01	1,2-Dichloroethane-d4	50	51.3	103	70 (63)	130 (155)
		Dibromofluoromethane	50	51.0	102	70 (70)	130 (134)
		Toluene-d8	50	51.1	102	70 (74)	130 (123)
		4-Bromofluorobenzene	50	48.5	97	70 (38)	130 (136)

