

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q2003 SAS No.: Q2003 SDG NO.: Q2003
 Lab File ID: VY022277.D Date Analyzed: 05/16/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:24
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	144477	13.346				
UPPER LIMIT	288954	13.846				
LOWER LIMIT	72238.5	12.846				
EPA SAMPLE NO.						
VOC	10942 *	13.35				
VY0516SBL01	151389	13.35				
VY0516SBS01	141210	13.35				
VY0516SBSD01	130982	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.