



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ardmore Chemical **SDG No.:** Q2006
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q2006 **SAS No.:** Q2006
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	11.1	12.0	92	65 - 135	P	05/14/2025	13:43	LB135778
	Zinc	42.2	40.0	106	65 - 135	P	05/14/2025	13:43	LB135778
CRI01	Lead	12.3	12.0	102	65 - 135	P	05/15/2025	13:51	LB135794
	Zinc	42.5	40.0	106	65 - 135	P	05/15/2025	13:51	LB135794
CRA	Mercury	0.059	0.05	118	60 - 140	CV	05/22/2025	14:37	LB135881