

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Ardmore Chemical **SDG No.:** Q2006
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** Q2006 **SAS No.:** Q2006
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	0.26			-12	12	05/14/2025	14:09	LB135778
	Zinc	3.75			-40	40	05/14/2025	14:09	LB135778
ICSAB01	Lead	46.1	49.0	94	37	61	05/14/2025	14:16	LB135778
	Zinc	981	952	103	809	1095	05/14/2025	14:16	LB135778
ICSA	Lead	4.80			-12	12	05/14/2025	14:43	LB135778
	Zinc	21.1			-40	40	05/14/2025	14:43	LB135778
ICSAB	Lead	23.8	49.0	49	37	61	05/14/2025	14:47	LB135778
	Zinc	1000	952	105	809	1095	05/14/2025	14:47	LB135778
ICSA01	Lead	1.30			-12	12	05/15/2025	14:01	LB135794
	Zinc	3.33			-40	40	05/15/2025	14:01	LB135794
ICSAB01	Lead	45.4	49.0	93	37	61	05/15/2025	14:05	LB135794
	Zinc	965	952	101	809	1095	05/15/2025	14:05	LB135794
ICSA	Lead	-2.62			-12	12	05/15/2025	14:09	LB135794
	Zinc	15.7			-40	40	05/15/2025	14:09	LB135794
ICSAB	Lead	65.2	49.0	133	37	61	05/15/2025	14:14	LB135794
	Zinc	970	952	102	809	1095	05/15/2025	14:14	LB135794