

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Ardmore Chemical</u>	level: <u>low</u>	sdg no.: <u>Q2006</u>
contract: <u>ARDM01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2006</u> sas no.: <u>Q2006</u>
matrix: <u>Water</u>	sample id: <u>Q2006-02</u>	client id: <u>EFF-WWMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2006-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	563		1.77	J	500	112		P
Mercury	ug/L	75 - 125	3.26		0.20	U	4.0	82		CV
Zinc	ug/L	75 - 125	356		231		100	125		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Ardmore Chemical **level:** low **sdg no.:** Q2006
contract: ARDM01 **lab code:** CHEM **case no.:** Q2006 **sas no.:** Q2006
matrix: Water **sample id:** Q2006-02 **client id:** EFF-WWMSD
Percent Solids for Sample: NA **Spiked ID:** Q2006-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	529		1.77	J	500	105		P
Mercury	ug/L	75 - 125	3.02		0.20	U	4.0	76		CV
Zinc	ug/L	75 - 125	275		231		100	44	N	P