

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical

SDG No.: Q2006

Contract: ARDM01

Lab Code: CHEM

Case No.: Q2006

SAS No.: Q2006

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167985BS							
Lead	ug/L	500	459		92	85 - 115	P
Zinc	ug/L	100	92.7		93	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Ardmore Chemical

SDG No.: Q2006

Contract: ARDM01

Lab Code: CHEM

Case No.: Q2006

SAS No.: Q2006

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168130BS Mercury	ug/L	4.0	3.82		96	85 - 115	CV