

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
Lab Code: CHEM Case No.: Q2006 SAS No.: Q2006 SDG No.: Q2006
Instrument ID: BNA_P Calibration Date(s): 05/13/2025 05/13/2025
Calibration Time(s): 10:41 15:26

LAB FILE ID:		RRF2.5 = BP024614.D			RRF005 = BP024615.D		RRF010 = BP024616.D	
		RRF020 = BP024617.D			RRF040 = BP024618.D		RRF050 = BP024619.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
n-Nitrosodimethylamine		0.536	0.507	0.527	0.517	0.580	0.538	4.6
2-Fluorophenol		1.134	1.096	1.162	1.105	1.272	1.169	5.5
Phenol-d6		1.301	1.344	1.486	1.472	1.671	1.492	9.1
Phenol		1.373	1.401	1.526	1.509	1.707	1.540	8.2
bis(2-Chloroethyl) ether		1.317	1.170	1.257	1.213	1.357	1.266	5.0
2-Chlorophenol		1.200	1.204	1.311	1.303	1.465	1.327	7.8
2,2-oxybis(1-Chloropropane)		1.543	1.445	1.520	1.467	1.617	1.516	4.2
n-Nitroso-di-n-propylamine	0.880	0.977	0.967	1.086	1.048	1.142	1.032	8.6
Nitrobenzene-d5		0.381	0.376	0.401	0.382	0.429	0.396	4.7
Hexachloroethane		0.606	0.591	0.582	0.551	0.623	0.589	4.0
Nitrobenzene		0.333	0.340	0.355	0.337	0.373	0.348	4.0
Isophorone		0.644	0.640	0.695	0.673	0.753	0.687	5.9
2-Nitrophenol		0.135	0.143	0.168	0.172	0.196	0.171	14.1
2,4-Dimethylphenol		0.273	0.279	0.300	0.298	0.329	0.301	6.6
bis(2-Chloroethoxy) methane		0.403	0.391	0.411	0.399	0.443	0.410	4.1
2,4-Dichlorophenol		0.250	0.258	0.290	0.291	0.327	0.292	10.0
1,2,4-Trichlorobenzene		0.346	0.322	0.327	0.306	0.347	0.329	4.3
Naphthalene		1.064	1.026	1.049	0.980	1.091	1.036	3.5
Hexachlorobutadiene		0.231	0.214	0.211	0.198	0.222	0.213	5.1
4-Chloro-3-methylphenol		0.316	0.325	0.358	0.356	0.391	0.356	7.8
Hexachlorocyclopentadiene		0.250	0.276	0.350	0.356	0.420	0.352	19.0
2,4,6-Trichlorophenol		0.305	0.341	0.375	0.371	0.428	0.376	11.3
2-Fluorobiphenyl		1.601	1.498	1.553	1.432	1.602	1.527	4.1
2-Chloronaphthalene		1.147	1.106	1.131	1.065	1.196	1.127	3.6
Dimethylphthalate		1.570	1.527	1.518	1.450	1.611	1.524	3.7
Acenaphthylene		1.845	1.818	1.920	1.804	2.029	1.886	4.2
2,6-Dinitrotoluene		0.299	0.308	0.326	0.312	0.354	0.322	5.8
Acenaphthene		1.105	1.054	1.094	1.028	1.159	1.084	4.0
2,4-Dinitrophenol			0.098	0.143	0.172	0.201	0.169	24.8
4-Nitrophenol			0.159	0.222	0.257	0.299	0.250	20.8
2,4-Dinitrotoluene		0.400	0.422	0.456	0.454	0.508	0.454	7.9
Diethylphthalate		1.593	1.542	1.537	1.444	1.639	1.541	4.4
4-Chlorophenyl-phenylether		0.752	0.697	0.700	0.656	0.741	0.703	5.0
Fluorene		1.466	1.415	1.410	1.336	1.494	1.410	4.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
Lab Code: CHEM Case No.: Q2006 SAS No.: Q2006 SDG No.: Q2006
Instrument ID: BNA_P Calibration Date(s): 05/13/2025 05/13/2025
Calibration Time(s): 10:41 15:26

LAB FILE ID:		RRF2.5 = BP024614.D			RRF005 = BP024615.D		RRF010 = BP024616.D	
		RRF020 = BP024617.D			RRF040 = BP024618.D		RRF050 = BP024619.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
4,6-Dinitro-2-methylphenol			0.094	0.120	0.129	0.149	0.129	15.6
n-Nitrosodiphenylamine		0.603	0.589	0.627	0.593	0.667	0.614	4.4
Azobenzene		1.327	1.299	1.356	1.250	1.407	1.318	4.4
2,4,6-Tribromophenol		0.324	0.321	0.335	0.323	0.369	0.339	5.5
4-Bromophenyl-phenylether		0.231	0.223	0.237	0.226	0.252	0.236	4.3
Hexachlorobenzene		0.303	0.288	0.298	0.282	0.321	0.299	4.2
Pentachlorophenol		0.124	0.137	0.163	0.172	0.198	0.168	17.1
Phenanthrene		1.144	1.085	1.125	1.051	1.170	1.112	3.7
Anthracene		1.097	1.064	1.142	1.075	1.205	1.119	4.3
Di-n-butylphthalate		1.182	1.218	1.332	1.260	1.447	1.302	7.0
Fluoranthene		1.271	1.254	1.314	1.250	1.399	1.300	4.2
Benzidine			0.362	0.543	0.560	0.657	0.545	18.2
Pyrene		1.184	1.137	1.197	1.153	1.273	1.198	3.8
Terphenyl-d14		1.192	1.113	1.165	1.125	1.226	1.166	3.3
Butylbenzylphthalate		0.458	0.465	0.537	0.527	0.604	0.533	10.3
3,3-Dichlorobenzidine		0.357	0.392	0.474	0.477	0.539	0.466	14.5
Benzo(a)anthracene		1.251	1.218	1.253	1.208	1.336	1.256	3.4
Chrysene		1.214	1.180	1.185	1.131	1.255	1.190	3.2
Bis(2-ethylhexyl)phthalate		0.673	0.683	0.801	0.766	0.884	0.783	10.3
Di-n-octyl phthalate		1.048	1.116	1.272	1.280	1.458	1.281	12.0
Benzo(b)fluoranthene		1.028	1.070	1.152	1.103	1.288	1.145	7.6
Benzo(k)fluoranthene		1.147	1.106	1.182	1.140	1.258	1.180	4.6
Benzo(a)pyrene		0.995	1.001	1.112	1.064	1.221	1.099	7.7
Indeno(1,2,3-cd)pyrene		1.341	1.357	1.447	1.424	1.608	1.460	6.7
Dibenzo(a,h)anthracene		1.068	1.066	1.186	1.167	1.321	1.188	8.2
Benzo(g,h,i)perylene		1.101	1.106	1.167	1.147	1.294	1.181	6.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1