

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ARDM01

Lab Code: CHEM Case No.: Q2006 SAS No.: Q2006 SDG No.: Q2006

Instrument ID: MSVOA_N Calibration Date/Time: 05/16/2025 11:37

Lab File ID: VN086661.D Init. Calib. Date(s): 04/23/2025 04/23/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:22 13:17

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acrolein	100.000	82.17	17.834	20