

Report of Analysis

Client:	Ardmore Chemical		Date Collected:		
Project:	PVSC Monthly 2025		Date Received:		
Client Sample ID:	VN0516WBSD01		SDG No.:	Q2006	
Lab Sample ID:	VN0516WBSD01		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-PP	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086663.D	1		05/16/25 12:46	VN051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
74-87-3	Chloromethane	18.4		0.64	5.00	ug/L
75-01-4	Vinyl Chloride	17.7		0.83	5.00	ug/L
74-83-9	Bromomethane	20.4		0.80	5.00	ug/L
75-00-3	Chloroethane	18.2		2.30	5.00	ug/L
75-69-4	Trichlorofluoromethane	19.5		0.80	5.00	ug/L
75-35-4	1,1-Dichloroethene	19.6		0.76	5.00	ug/L
107-02-8	Acrolein	130		6.60	25.0	ug/L
107-13-1	Acrylonitrile	110		2.80	25.0	ug/L
75-09-2	Methylene Chloride	20.4		0.86	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.1		0.82	5.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.68	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.6		0.74	5.00	ug/L
67-66-3	Chloroform	19.8		0.55	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.63	5.00	ug/L
71-43-2	Benzene	20.1		0.45	5.00	ug/L
107-06-2	1,2-Dichloroethane	20.4		0.50	5.00	ug/L
79-01-6	Trichloroethene	19.6		0.49	5.00	ug/L
78-87-5	1,2-Dichloropropane	19.9		0.46	5.00	ug/L
75-27-4	Bromodichloromethane	20.2		0.64	5.00	ug/L
108-88-3	Toluene	19.7		0.46	5.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.2		0.72	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.3		0.67	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.1		0.45	5.00	ug/L
110-75-8	2-Chloroethyl vinyl ether	93.8		4.60	25.0	ug/L
124-48-1	Dibromochloromethane	20.1		0.66	5.00	ug/L
127-18-4	Tetrachloroethene	19.9		0.84	5.00	ug/L
108-90-7	Chlorobenzene	20.1		0.47	5.00	ug/L
100-41-4	Ethyl Benzene	19.2		0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	38.8		1.30	10.0	ug/L
95-47-6	o-Xylene	19.5		0.67	5.00	ug/L

Report of Analysis

Client:	Ardmore Chemical		Date Collected:		
Project:	PVSC Monthly 2025		Date Received:		
Client Sample ID:	VN0516WBSD01		SDG No.:	Q2006	
Lab Sample ID:	VN0516WBSD01		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-PP	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086663.D	1		05/16/25 12:46	VN051625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
75-25-2	Bromoform	20.2		0.94	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	22.9		0.44	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.2		0.67	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.2		0.81	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.0		0.67	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	29.2		91 - 110	97%	SPK: 30
2037-26-5	Toluene-d8	29.5		91 - 112	98%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.9		63 - 112	96%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	28900	7.806			
540-36-3	1,4-Difluorobenzene	168000	9.1			
3114-55-4	Chlorobenzene-d5	156000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products