

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: Q2006 SAS No.: Q2006 SDG NO.: Q2006
 Lab File ID: VN086661.D Date Analyzed: 05/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	29831	7.81	175700	9.10	163950	11.87
UPPER LIMIT	59662	8.306	351400	9.6	327900	12.365
LOWER LIMIT	14915.5	7.306	87850	8.6	81975	11.365
EPA SAMPLE NO.						
EFF-WW	28982	7.81	166943	9.10	151707	11.87
VN0516WBL01	29934	7.81	171805	9.10	155509	11.87
VN0516WBS01	28843	7.81	173361	9.10	161573	11.87
VN0516WBSD01	28926	7.81	168023	9.10	156348	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.