



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger Site Princeton NJ 2025**

**Project # N/A**

**Order ID # Q2008**

**Test Name: Flash Point,pH**

### **A. Number of Samples and Date of Receipt:**

1 Water sample was received on 05/09/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Diesel Range Organics, Flash Point, Gasoline Range Organics, Mercury, Metals ICP-TAL, METALS-TAL, pH, SVOC-TCL BNA -20 and VOC-TCLVOA-10. This data package contains results for Flash Point,pH.

### **C. Analytical Techniques:**

The analysis of Flash Point was based on method 1010B and The analysis of pH was based on method 9040C.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for IDW-AQ-DRUM-633-05092025 of pH, for IDW-AQ-DRUM-633-05092025 of pH as samples were receive out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_