## DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name :		Alliance Technical Group LLC	Client :	JACOBS Engineering Group, Inc.					
Project Location :		Princeton Junction NJ	Project Number :	D386822					
Laboratory Sample ID(s): Q2008 Sampling Date(s): 05/09/2025									
List DKQP Methods Used (e.g., 8260,8270, et Cetra) ,1010B,6010D,7470A,8015D,8260D,8270E,9040									
1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?					Yes		No	
1A	Were the method specified handling, preservation, and holding time requirements met?					Yes	$\checkmark$	No	
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)					Yes		No	✓ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?				V	Yes		No	
3	Were samples received at an appropriate temperature (4±2° C)?				$\overline{\mathbf{A}}$	Yes		No	□ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?					Yes	$\checkmark$	No	
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt?				V	Yes		No	
	b)Were these	reporting limits met?			V	Yes		No	□ N/A
6	results report	ytical method referenced in this labored for all constituents identified in the DKQP documents and/or site-species.	ne method-specific analyte		V	Yes		No	
7	Are project-sp	ecific matrix spikes and/or laborator	y duplicates included in thi	s data set?		Yes	$\overline{\checkmark}$	No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."