

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135743

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/12/2025	15:05
2	CAL2	1	Water	NA	NA	20.2	7.00	05/12/2025	15:06
3	CAL3	1	Water	NA	NA	20.2	10.02	05/12/2025	15:10
4	ICV	1	Water	NA	NA	20.2	7.01	05/12/2025	15:15
5	CCV1	1	Water	NA	NA	20.1	2.01	05/12/2025	15:16
6	Q2008-01	1	Water	NA	NA	20.7	2.06	05/12/2025	15:30
7	Q2008-01DUP	1	Water	NA	NA	20.8	2.07	05/12/2025	15:31
8	CCV2	1	Water	NA	NA	20.2	12.02	05/12/2025	15:33

VB 135743

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w q2008 WorkList ID : 189458 Department : Wet-Chemistry Date : 05-12-2025 12:57:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2008-01	E IDW-AQ-DRUM-633-05092025	Water	pH	Cool 4 deg C	JACO05	L41	05/09/2025	9040C

Date/Time 05/12/25 13:10
Raw Sample Received by: SR WOC
Raw Sample Relinquished by: JTCsm

Date/Time 05-12-25 16:30
Raw Sample Received by: JTCsm
Raw Sample Relinquished by: SR WOC