

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2008

Project: Former Schlumberger Site Princeton NJ 2025

RunNo.: LB135743

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	05/12/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	05/12/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	05/12/2025

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2008

Project: Former Schlumberger Site Princeton NJ 2025

RunNo.: LB135765

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	83.9	81	104	78-84	05/14/2025

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2008

Project: Former Schlumberger Site Princeton NJ 2025

RunNo.: LB135801

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	05/16/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	05/16/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	05/16/2025

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2008

Project: Former Schlumberger Site Princeton NJ 2025

RunNo.: LB135801

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------