

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2008
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2008 **SAS No.:** Q2008

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167955BS							
Aluminum	ug/L	1000	914		91	80 - 120	P
Antimony	ug/L	400	382		96	80 - 120	P
Arsenic	ug/L	400	382		96	80 - 120	P
Barium	ug/L	100	107		107	80 - 120	P
Beryllium	ug/L	100	96.5		96	80 - 120	P
Cadmium	ug/L	100	97.5		98	80 - 120	P
Calcium	ug/L	500	461	J	92	80 - 120	P
Chromium	ug/L	200	190		95	80 - 120	P
Cobalt	ug/L	100	94.5		94	80 - 120	P
Copper	ug/L	150	145		97	80 - 120	P
Iron	ug/L	1500	1400		93	80 - 120	P
Lead	ug/L	500	473		95	80 - 120	P
Magnesium	ug/L	1000	872	J	87	80 - 120	P
Manganese	ug/L	100	92.3		92	80 - 120	P
Nickel	ug/L	250	240		96	80 - 120	P
Potassium	ug/L	5000	4550		91	80 - 120	P
Selenium	ug/L	1000	973		97	80 - 120	P
Silver	ug/L	37.5	34.3		92	80 - 120	P
Sodium	ug/L	1500	1330		89	80 - 120	P
Thallium	ug/L	1000	968		97	80 - 120	P
Vanadium	ug/L	150	139		93	80 - 120	P
Zinc	ug/L	100	94.6		95	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2008

Contract: JACO05

Lab Code: CHEM

Case No.: Q2008

SAS No.: Q2008

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167980BS Mercury	ug/L	4.0	3.71		93	82 - 119	CV