

**WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE**

**Lab Name:** Chemtech **Client:** JACOBS Engineering Group, Inc.  
**Lab Code:** CHEM **Cas No:** Q2008 **SAS No :** Q2008 **SDG No:** Q2008  
**Matrix Spike - EPA Sample No :** PB167981BSD **Datafile:** FF015832.D

| COMPOUND | SPIKE<br>ADDED<br>ug/L | CONCENTRATION<br>ug/L | LCS/LCSD<br>CONCENTRATION<br>ug/L | % REC | QC LIMITS |
|----------|------------------------|-----------------------|-----------------------------------|-------|-----------|
| DRO      | 200                    | 0                     | 203                               | 102   | 78-117    |

LCS/LCSD % Recovery RPD : 2.0