

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB042325

Review By	yogesh	Review On	4/23/2025 1:56:12 PM
Supervise By	mohammad	Supervise On	4/26/2025 2:19:22 AM
SubDirectory	FB042325	HP Acquire Method	HP Processing Method
			FB042325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24485,PP24486,PP24487,PP24488,PP24489		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24490		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031638.D	23 Apr 2025 12:11		YP/AJ	Ok
2	10 GRO STD		FB031639.D	23 Apr 2025 12:38		YP/AJ	Ok
3	20 GRO STD		FB031640.D	23 Apr 2025 13:05		YP/AJ	Ok
4	50 GRO STD		FB031641.D	23 Apr 2025 13:32		YP/AJ	Ok,M
5	100 GRO STD		FB031642.D	23 Apr 2025 14:00		YP/AJ	Ok,M
6	FB042325GROICV		FB031643.D	23 Apr 2025 14:39		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB051225

Review By	yogesh	Review On	5/12/2025 1:58:29 PM
Supervise By	mohammad	Supervise On	5/14/2025 5:13:59 AM
SubDirectory	FB051225	HP Acquire Method	HP Processing Method
			FB042325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24485,PP24486,PP24487,PP24488,PP24489
CCC Internal Standard/PEM	PP24563,PP24564,PP24565,PP24566
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24490

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031694.D	12 May 2025 9:40		YP/AJ	Ok,M
2	VBF0512W1		FB031695.D	12 May 2025 10:20		YP/AJ	Ok
3	BSF0512W1		FB031696.D	12 May 2025 11:16		YP/AJ	Ok
4	Q2008-01		FB031697.D	12 May 2025 11:44	Vial-A	YP/AJ	Ok
5	BSF0512W2		FB031698.D	12 May 2025 12:55		YP/AJ	Ok,M
6	20 PPB GRO STD		FB031699.D	12 May 2025 13:53		YP/AJ	Ok,M
7	VBF0512S1		FB031700.D	12 May 2025 14:37	end ccc missing	YP/AJ	Not Ok
8	VBF0512S2		FB031701.D	12 May 2025 15:23	end ccc missing	YP/AJ	Not Ok
9	BSF0512S1		FB031702.D	12 May 2025 16:41	end ccc missing	YP/AJ	Not Ok
10	Q1982-01		FB031703.D	12 May 2025 17:12	surrogate fail ,end ccc missing	YP/AJ	Not Ok

M : Manual Integration