

WATER GASOLINE RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATION

Lab Name: Chemtech **Client:** JACOBS Engineering Group, Inc.
Lab Code: CHEM **Cas No:** Q2008 **SAS No :** Q2008 **SDG No:** Q2008
Matrix Spike - EPA Sample No : BSF0512W2 **Datafile:** FB031698.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
GRO	180	0	185	103	50-150

LCS/LCSD % Recovery RPD : 12.6