

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q2008 SAS No.: Q2008 SDG No.: Q2008

Instrument ID: MSVOA_X Calibration Date/Time: 05/12/2025 09:59

Lab File ID: VX046125.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.765	0.819		7.06	20
Chloromethane	0.742	0.738	0.1	-0.54	20
Vinyl Chloride	0.691	0.666		-3.62	20
Bromomethane	0.320	0.303		-5.31	20
Chloroethane	0.369	0.371		0.54	20
Trichlorofluoromethane	1.021	1.041		1.96	20
1,1,2-Trichlorotrifluoroethane	0.632	0.622		-1.58	20
1,1-Dichloroethene	0.593	0.562		-5.23	20
Acetone	0.374	0.357		-4.55	20
Carbon Disulfide	1.406	1.418		0.85	20
Methyl tert-butyl Ether	2.079	2.008		-3.41	20
Methyl Acetate	0.867	0.843		-2.77	20
Methylene Chloride	0.716	0.643		-10.2	20
trans-1,2-Dichloroethene	0.596	0.570		-4.36	20
1,1-Dichloroethane	1.219	1.174	0.1	-3.69	20
Cyclohexane	1.111	1.042		-6.21	20
2-Butanone	0.543	0.506		-6.81	20
Carbon Tetrachloride	0.544	0.557		2.39	20
cis-1,2-Dichloroethene	0.718	0.686		-4.46	20
Bromochloromethane	0.587	0.557		-5.11	20
Chloroform	1.271	1.225		-3.62	20
1,1,1-Trichloroethane	1.101	1.086		-1.36	20
Methylcyclohexane	0.623	0.612		-1.77	20
Benzene	1.417	1.405		-0.85	20
1,2-Dichloroethane	0.612	0.610		-0.33	20
Trichloroethene	0.341	0.339		-0.59	20
1,2-Dichloropropane	0.352	0.360		2.27	20
Bromodichloromethane	0.547	0.578		5.67	20
4-Methyl-2-Pentanone	0.605	0.596		-1.49	20
Toluene	0.869	0.853		-1.84	20
t-1,3-Dichloropropene	0.487	0.519		6.57	20
cis-1,3-Dichloropropene	0.538	0.555		3.16	20
1,1,2-Trichloroethane	0.343	0.346		0.88	20
2-Hexanone	0.448	0.439		-2.01	20
Dibromochloromethane	0.376	0.412		9.57	20
1,2-Dibromoethane	0.356	0.351		-1.4	20
Tetrachloroethene	0.354	0.352		-0.56	20
Chlorobenzene	1.094	1.063	0.3	-2.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JACO05

Lab Code: CHEM Case No.: Q2008 SAS No.: Q2008 SDG No.: Q2008

Instrument ID: MSVOA_X Calibration Date/Time: 05/12/2025 09:59

Lab File ID: VX046125.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	1.941		0.62	20
m/p-Xylenes	0.706	0.710		0.57	20
o-Xylene	0.688	0.691		0.44	20
Styrene	1.127	1.178		4.53	20
Bromoform	0.281	0.302	0.1	7.47	20
Isopropylbenzene	3.893	3.916		0.59	20
1,1,2,2-Tetrachloroethane	1.364	1.243	0.3	-8.87	20
1,3-Dichlorobenzene	1.649	1.648		-0.06	20
1,4-Dichlorobenzene	1.684	1.688		0.24	20
1,2-Dichlorobenzene	1.655	1.616		-2.36	20
1,2-Dibromo-3-Chloropropane	0.302	0.298		-1.33	20
1,2,4-Trichlorobenzene	0.951	0.964		1.37	20
1,2,3-Trichlorobenzene	0.981	0.982		0.1	20
1,2-Dichloroethane-d4	0.932	0.830		-10.94	20
Dibromofluoromethane	0.360	0.350		-2.78	20
Toluene-d8	1.246	1.176		-5.62	20
4-Bromofluorobenzene	0.478	0.462		-3.35	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.