

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q2008 SAS No.: Q2008 SDG NO.: Q2008
Lab File ID: VX046125.D Date Analyzed: 05/12/2025
Instrument ID: MSVOA_X Time Analyzed: 09:59
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	99352	5.54	165662	6.75	144793	10.05
UPPER LIMIT	198704	6.038	331324	7.251	289586	10.549
LOWER LIMIT	49676	5.038	82831	6.251	72396.5	9.549
EPA SAMPLE NO.						
IDW-AQ-DRUM-633-05092025	68762	5.54	138646	6.76	129122	10.05
VX0512WBL01	69374	5.54	138547	6.76	131027	10.05
VX0512WBS01	93419	5.54	162432	6.76	143385	10.05
VX0512WBSD01	90547	5.54	155274	6.76	138103	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.