

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2008 SAS No.: Q2008 SDG NO.: Q2008
 Lab File ID: VX046125.D Date Analyzed: 05/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69015	12.018				
UPPER LIMIT	138030	12.518				
LOWER LIMIT	34507.5	11.518				
EPA SAMPLE NO.						
IDW-AQ-DRUM-633-05092025	57082	12.02				
VX0512WBL01	55247	12.02				
VX0512WBS01	66250	12.02				
VX0512WBSD01	65121	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.