



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q2010			
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q2010	SAS No.:	Q2010	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB07	Mercury	0.20	+/-0.20	U	0.20	CV	05/13/2025	11:09	LB135751

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2010

Contract: CAMP02

Lab Code: CHEM

Case No.: Q2010

SAS No.: Q2010

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB17	Mercury	0.20	+/-0.20	U	0.20	CV	05/13/2025	11:14	LB135751
CCB18	Mercury	0.20	+/-0.20	U	0.20	CV	05/13/2025	11:46	LB135751
CCB19	Mercury	0.20	+/-0.20	U	0.20	CV	05/13/2025	12:19	LB135751
CCB20	Mercury	0.20	+/-0.20	U	0.20	CV	05/13/2025	12:46	LB135751
CCB21	Mercury	0.20	+/-0.20	U	0.20	CV	05/13/2025	13:19	LB135751

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith SDG No.: Q2010
Contract: CAMP02 Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	05/16/2025	13:05	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	13:05	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	13:05	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	13:05	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	13:05	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	13:05	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	13:05	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	13:05	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	13:05	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	13:05	LB135802
	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	13:05	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	13:05	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	13:05	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	13:05	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	13:05	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	13:05	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	13:05	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	13:05	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2010							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2010	SAS No.: Q2010						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	05/16/2025	13:39	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	13:39	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	13:39	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	13:39	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	13:39	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	13:39	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	13:39	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	13:39	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	13:39	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	13:39	LB135802
	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	13:39	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	13:39	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	13:39	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	13:39	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	13:39	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	13:39	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	13:39	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	13:39	LB135802
CCB02	Aluminum	100	+/-100	U	100	P	05/16/2025	14:40	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	14:40	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	14:40	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	14:40	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	14:40	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	14:40	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	14:40	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	14:40	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	14:40	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	14:40	LB135802
	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	14:40	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	14:40	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	14:40	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2010							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2010	SAS No.: Q2010						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	14:40	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	14:40	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	14:40	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	14:40	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	14:40	LB135802
CCB03	Aluminum	100	+/-100	U	100	P	05/16/2025	15:26	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	15:26	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	15:26	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	15:26	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	15:26	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	15:26	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	15:26	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	15:26	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	15:26	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	15:26	LB135802
	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	15:26	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	15:26	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	15:26	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	15:26	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	15:26	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	15:26	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	15:26	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	15:26	LB135802
CCB04	Aluminum	100	+/-100	U	100	P	05/16/2025	16:12	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	16:12	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	16:12	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	16:12	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	16:12	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	16:12	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	16:12	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	16:12	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	16:12	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	16:12	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2010							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2010	SAS No.: Q2010						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	16:12	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	16:12	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	16:12	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	16:12	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	16:12	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	16:12	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	16:12	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	16:12	LB135802
CCB05	Aluminum	100	+/-100	U	100	P	05/16/2025	16:58	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	16:58	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	16:58	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	16:58	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	16:58	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	16:58	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	16:58	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	16:58	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	16:58	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	16:58	LB135802
	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	16:58	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	16:58	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	16:58	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	16:58	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	16:58	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	16:58	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	16:58	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	16:58	LB135802
CCB06	Aluminum	100	+/-100	U	100	P	05/16/2025	17:32	LB135802
	Antimony	50.0	+/-50.0	U	50.0	P	05/16/2025	17:32	LB135802
	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	17:32	LB135802
	Beryllium	6.00	+/-6.00	U	6.00	P	05/16/2025	17:32	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	17:32	LB135802
	Calcium	2000	+/-2000	U	2000	P	05/16/2025	17:32	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2010							
Contract: CAMP02		Lab Code: CHEM		Case No.: Q2010		SAS No.: Q2010			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	17:32	LB135802
	Cobalt	30.0	+/-30.0	U	30.0	P	05/16/2025	17:32	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Iron	100	+/-100	U	100	P	05/16/2025	17:32	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	17:32	LB135802
	Magnesium	2000	+/-2000	U	2000	P	05/16/2025	17:32	LB135802
	Manganese	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	17:32	LB135802
	Potassium	2000	+/-2000	U	2000	P	05/16/2025	17:32	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	17:32	LB135802
	Sodium	2000	+/-2000	U	2000	P	05/16/2025	17:32	LB135802
	Thallium	40.0	+/-40.0	U	40.0	P	05/16/2025	17:32	LB135802
	Vanadium	40.0	+/-40.0	U	40.0	P	05/16/2025	17:32	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	17:32	LB135802