

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2010
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2010 **SAS No.:** Q2010

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167977BS							
Aluminum	mg/Kg	100	90.8		91	80 - 120	P
Antimony	mg/Kg	40.0	39.4		98	80 - 120	P
Arsenic	mg/Kg	40.0	38.5		96	80 - 120	P
Barium	mg/Kg	10.0	8.83		88	80 - 120	P
Beryllium	mg/Kg	10.0	9.29		93	80 - 120	P
Cadmium	mg/Kg	10.0	9.64		96	80 - 120	P
Calcium	mg/Kg	50.0	46.8	J	94	80 - 120	P
Chromium	mg/Kg	20.0	19.1		96	80 - 120	P
Cobalt	mg/Kg	10.0	9.37		94	80 - 120	P
Copper	mg/Kg	15.0	14.7		98	80 - 120	P
Iron	mg/Kg	150	141		94	80 - 120	P
Lead	mg/Kg	50.0	46.9		94	80 - 120	P
Magnesium	mg/Kg	100	90.4	J	90	80 - 120	P
Manganese	mg/Kg	10.0	9.15		92	80 - 120	P
Nickel	mg/Kg	25.0	23.6		94	80 - 120	P
Potassium	mg/Kg	500	457		91	80 - 120	P
Selenium	mg/Kg	100	98.4		98	80 - 120	P
Silver	mg/Kg	3.8	3.53		93	80 - 120	P
Sodium	mg/Kg	150	139		93	80 - 120	P
Thallium	mg/Kg	100	97.5		98	80 - 120	P
Vanadium	mg/Kg	15.0	13.8		92	80 - 120	P
Zinc	mg/Kg	10.0	9.69		97	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2010
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2010 **SAS No.:** Q2010

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167983BS Mercury	mg/Kg	0.24	0.22		90	80 - 120	CV