

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Date Collected:

Date Received:

Q2010

Decanted:

Fax: 908 789 8922

Report of Analysis

Client: CDM Smith

Project: South River WM Replacement

Client Sample ID: BSF0514S1 SDG No.:

Lab Sample ID:BSF0514S1Matrix:SOILAnalytical Method:8015D GRO% Solid:100

Sample Wt/Vol: 5 Units: g Final Vol: 5 mL

Soil Aliquot Vol: uL Test: Gasoline Range Organics

Extraction Type: Injection Volume:

GPC Factor: PH:

Prep Method:

File ID/Qc Batch: Dilution: Date Analyzed Prep Batch ID

FB031738.D 1 05/14/25 11:31 FB051425

Qualifier MDL LOQ / CRQL Units(Dry Weight) **CAS Number Parameter** Conc. **TARGETS GRO GRO** 184 8.00 45.0 ug/kg **SURROGATES** 98-08-8 Alpha, Alpha, Alpha-Trifluoroto 23.8 50 - 150 119% SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit