

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030180.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.132	9.033	9.233	683.550	712.500	-4.1
2,4,5-TP(Silvex)	8.850	8.750	8.950	681.540	712.500	-4.3
2,4-D	8.009	7.910	8.110	665.970	705.000	-5.5
2,4-DB	9.690	9.589	9.789	709.650	712.500	-0.4
2,4-DCAA	6.933	6.834	7.034	709.920	750.000	-5.3
DICAMBA	7.109	7.010	7.210	674.500	705.000	-4.3
DICHLORPROP	7.789	7.689	7.889	663.920	705.000	-5.8
Dinoseb	10.850	10.751	10.951	667.910	705.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030180.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.925	9.826	10.026	685.760	712.500	-3.8
2,4,5-TP(Silvex)	9.521	9.423	9.623	691.060	712.500	-3.0
2,4-D	8.643	8.544	8.744	682.250	705.000	-3.2
2,4-DB	10.483	10.385	10.585	647.540	712.500	-9.1
2,4-DCAA	7.453	7.354	7.554	717.910	750.000	-4.3
DICAMBA	7.638	7.539	7.739	686.280	705.000	-2.7
DICHLORPROP	8.331	8.232	8.432	661.570	705.000	-6.2
Dinoseb	10.856	10.758	10.958	670.800	705.000	-4.9