

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030190.D Time Analyzed: 17:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.132	9.033	9.233	698.460	712.500	-2.0
2,4,5-TP(Silvex)	8.851	8.750	8.950	696.850	712.500	-2.2
2,4-D	8.010	7.910	8.110	679.920	705.000	-3.6
2,4-DB	9.690	9.589	9.789	727.790	712.500	2.1
2,4-DCAA	6.933	6.834	7.034	724.480	750.000	-3.4
DICAMBA	7.110	7.010	7.210	689.130	705.000	-2.3
DICHLORPROP	7.790	7.689	7.889	677.360	705.000	-3.9
Dinoseb	10.851	10.751	10.951	683.150	705.000	-3.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030190.D Time Analyzed: 17:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.925	9.826	10.026	702.590	712.500	-1.4
2,4,5-TP(Silvex)	9.521	9.423	9.623	706.930	712.500	-0.8
2,4-D	8.643	8.544	8.744	697.370	705.000	-1.1
2,4-DB	10.483	10.385	10.585	660.760	712.500	-7.3
2,4-DCAA	7.453	7.354	7.554	734.300	750.000	-2.1
DICAMBA	7.639	7.539	7.739	704.930	705.000	0.0
DICHLORPROP	8.331	8.232	8.432	679.360	705.000	-3.6
Dinoseb	10.857	10.758	10.958	688.750	705.000	-2.3