

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030236.D Time Analyzed: 16:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.131	9.033	9.233	707.730	712.500	-0.7
2,4,5-TP(Silvex)	8.849	8.750	8.950	707.610	712.500	-0.7
2,4-D	8.008	7.910	8.110	690.730	705.000	-2.0
2,4-DB	9.689	9.589	9.789	753.460	712.500	5.7
2,4-DCAA	6.932	6.834	7.034	734.390	750.000	-2.1
DICAMBA	7.109	7.010	7.210	700.170	705.000	-0.7
DICHLORPROP	7.789	7.689	7.889	688.270	705.000	-2.4
Dinoseb	10.850	10.751	10.951	681.630	705.000	-3.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030236.D Time Analyzed: 16:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.923	9.826	10.026	719.900	712.500	1.0
2,4,5-TP(Silvex)	9.519	9.423	9.623	722.990	712.500	1.5
2,4-D	8.641	8.544	8.744	710.880	705.000	0.8
2,4-DB	10.481	10.385	10.585	666.280	712.500	-6.5
2,4-DCAA	7.451	7.354	7.554	743.230	750.000	-0.9
DICAMBA	7.636	7.539	7.739	718.480	705.000	1.9
DICHLORPROP	8.329	8.232	8.432	689.300	705.000	-2.2
Dinoseb	10.855	10.758	10.958	701.990	705.000	-0.4