

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030221.D Time Analyzed: 08:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.132	9.033	9.233	715.310	712.500	0.4
2,4,5-TP(Silvex)	8.850	8.750	8.950	713.830	712.500	0.2
2,4-D	8.009	7.910	8.110	697.060	705.000	-1.1
2,4-DB	9.689	9.589	9.789	741.130	712.500	4.0
2,4-DCAA	6.933	6.834	7.034	742.920	750.000	-0.9
DICAMBA	7.109	7.010	7.210	706.540	705.000	0.2
DICHLORPROP	7.789	7.689	7.889	696.890	705.000	-1.2
Dinoseb	10.850	10.751	10.951	701.370	705.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030221.D Time Analyzed: 08:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.924	9.826	10.026	716.400	712.500	0.5
2,4,5-TP(Silvex)	9.521	9.423	9.623	723.050	712.500	1.5
2,4-D	8.643	8.544	8.744	713.300	705.000	1.2
2,4-DB	10.482	10.385	10.585	640.930	712.500	-10.0
2,4-DCAA	7.452	7.354	7.554	749.890	750.000	0.0
DICAMBA	7.638	7.539	7.739	720.210	705.000	2.2
DICHLORPROP	8.330	8.232	8.432	697.900	705.000	-1.0
Dinoseb	10.855	10.758	10.958	705.820	705.000	0.1