

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2010

Analytical Method: 8151A

Client: CDM Smith

Datafile : PS030216.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB167976BS (Column 1)	DICAMBA	166.5	166	ug/Kg	100				72	129
	DICHLORPROP	166.5	166	ug/Kg	100				77	135
	2,4-D	166.5	166	ug/Kg	100				65	144
	2,4,5-TP(Silvex)	166.5	170	ug/Kg	102				74	146
	2,4,5-T	166.5	170	ug/Kg	102				77	134
	2,4-DB	166.5	174	ug/Kg	105				72	122
	Dinoseb	166.5	165	ug/Kg	99				74	132
PB167976BS (Column 2)	DICAMBA	166.5	164	ug/Kg	98				72	129
	DICHLORPROP	166.5	163	ug/Kg	98				77	135
	2,4-D	166.5	167	ug/Kg	100				65	144
	2,4,5-TP(Silvex)	166.5	167	ug/Kg	100				74	146
	2,4,5-T	166.5	166	ug/Kg	100				77	134
	2,4-DB	166.5	147	ug/Kg	88				72	122
	Dinoseb	166.5	162	ug/Kg	97				74	132