

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088537.D Time Analyzed: 19:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.606	6.806	55.770	50.000	11.5
4,4'-DDE	6.196	6.097	6.297	52.660	50.000	5.3
4,4'-DDT	7.021	6.922	7.122	52.310	50.000	4.6
Aldrin	5.273	5.173	5.373	56.150	50.000	12.3
alpha-BHC	4.001	3.901	4.101	56.370	50.000	12.7
alpha-Chlordane	6.028	5.929	6.129	57.060	50.000	14.1
beta-BHC	4.516	4.416	4.616	54.700	50.000	9.4
Decachlorobiphenyl	9.072	8.975	9.175	48.530	50.000	-2.9
delta-BHC	4.765	4.665	4.865	55.790	50.000	11.6
Dieldrin	6.348	6.249	6.449	55.340	50.000	10.7
Endosulfan I	6.075	5.976	6.176	56.160	50.000	12.3
Endosulfan II	6.786	6.687	6.887	53.180	50.000	6.4
Endosulfan sulfate	7.150	7.050	7.250	53.210	50.000	6.4
Endrin	6.575	6.476	6.676	53.640	50.000	7.3
Endrin aldehyde	6.915	6.816	7.016	52.180	50.000	4.4
Endrin ketone	7.630	7.531	7.731	52.880	50.000	5.8
gamma-BHC (Lindane)	4.331	4.232	4.432	55.450	50.000	10.9
gamma-Chlordane	5.945	5.847	6.047	54.760	50.000	9.5
Heptachlor	4.931	4.831	5.031	56.180	50.000	12.4
Heptachlor epoxide	5.691	5.592	5.792	55.000	50.000	10.0
Methoxychlor	7.493	7.395	7.595	50.660	50.000	1.3
Tetrachloro-m-xylene	3.551	3.452	3.652	53.660	50.000	7.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: PSTDCCC050 Data File : PD088537.D Time Analyzed: 19:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	47.580	50.000	-4.8
4,4'-DDE	5.375	5.280	5.480	47.480	50.000	-5.0
4,4'-DDT	6.186	6.088	6.288	46.850	50.000	-6.3
Aldrin	4.369	4.273	4.473	48.520	50.000	-3.0
alpha-BHC	3.395	3.296	3.496	48.820	50.000	-2.4
alpha-Chlordane	5.192	5.094	5.294	47.350	50.000	-5.3
beta-BHC	4.027	3.928	4.128	47.720	50.000	-4.6
Decachlorobiphenyl	8.074	7.977	8.177	40.950	50.000	-18.1
delta-BHC	4.263	4.165	4.365	48.730	50.000	-2.5
Dieldrin	5.515	5.417	5.617	47.510	50.000	-5.0
Endosulfan I	5.248	5.151	5.351	40.920	50.000	-18.2
Endosulfan II	6.082	5.984	6.184	47.150	50.000	-5.7
Endosulfan sulfate	6.484	6.386	6.586	45.830	50.000	-8.3
Endrin	5.791	5.693	5.893	47.710	50.000	-4.6
Endrin aldehyde	6.260	6.163	6.363	45.360	50.000	-9.3
Endrin ketone	6.993	6.895	7.095	45.410	50.000	-9.2
gamma-BHC (Lindane)	3.731	3.633	3.833	48.050	50.000	-3.9
gamma-Chlordane	5.128	5.030	5.230	47.690	50.000	-4.6
Heptachlor	4.085	3.986	4.186	48.050	50.000	-3.9
Heptachlor epoxide	4.874	4.777	4.977	47.860	50.000	-4.3
Methoxychlor	6.756	6.659	6.859	45.410	50.000	-9.2
Tetrachloro-m-xylene	2.882	2.783	2.983	48.160	50.000	-3.7