

Surrogate Summary

SDG No.: Q2010

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088121.D	PIBLK-PD088121.D	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	22.7	113		43	140
		Tetrachloro-m-xylene	2	20	20.9	104		77	126
I.BLK-PD088523.D	PIBLK-PD088523.D	Decachlorobiphenyl	1	20	19.3	97		43	140
		Tetrachloro-m-xylene	1	20	18.3	92		77	126
		Decachlorobiphenyl	2	20	17.4	87		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126
PB167959BL	PB167959BL	Decachlorobiphenyl	1	20	20.1	100		20	144
		Tetrachloro-m-xylene	1	20	19.3	97		19	148
		Decachlorobiphenyl	2	20	18.1	90		20	144
		Tetrachloro-m-xylene	2	20	19.1	96		19	148
PB167959BS	PB167959BS	Decachlorobiphenyl	1	20	22.7	114		20	144
		Tetrachloro-m-xylene	1	20	21.6	108		19	148
		Decachlorobiphenyl	2	20	20.6	103		20	144
		Tetrachloro-m-xylene	2	20	21.3	106		19	148
Q2010-01	TP-10	Decachlorobiphenyl	1	20	11.7	58		20	144
		Tetrachloro-m-xylene	1	20	12.8	64		19	148
		Decachlorobiphenyl	2	20	9.76	49		20	144
		Tetrachloro-m-xylene	2	20	12.1	61		19	148
Q2010-02	TP-13	Decachlorobiphenyl	1	20	15.2	76		20	144
		Tetrachloro-m-xylene	1	20	17.6	88		19	148
		Decachlorobiphenyl	2	20	13.7	68		20	144
		Tetrachloro-m-xylene	2	20	16.9	85		19	148
Q2010-03	TP-14	Decachlorobiphenyl	1	20	16.5	83		20	144
		Tetrachloro-m-xylene	1	20	18.3	91		19	148
		Decachlorobiphenyl	2	20	13.5	68		20	144
		Tetrachloro-m-xylene	2	20	17.5	88		19	148
Q2010-04	TP-17	Decachlorobiphenyl	1	20	14.8	74		20	144
		Tetrachloro-m-xylene	1	20	18.4	92		19	148
		Decachlorobiphenyl	2	20	13.1	65		20	144
		Tetrachloro-m-xylene	2	20	17.7	88		19	148
I.BLK-PD088535.D	PIBLK-PD088535.D	Decachlorobiphenyl	1	20	18.0	90		43	140
		Tetrachloro-m-xylene	1	20	17.4	87		77	126
		Decachlorobiphenyl	2	20	14.8	74		43	140
		Tetrachloro-m-xylene	2	20	16.4	82		77	126
I.BLK-PD088551.D	PIBLK-PD088551.D	Decachlorobiphenyl	1	20	20.9	104		43	140
		Tetrachloro-m-xylene	1	20	20.3	101		77	126
		Decachlorobiphenyl	2	20	18.6	93		43	140
		Tetrachloro-m-xylene	2	20	19.1	95		77	126
Q1984-01MS	OU4-PCS-TC-33-050725MS	Decachlorobiphenyl	1	20	21.1	106		20	144

Surrogate Summary

SDG No.: Q2010

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1984-01MS	OU4-PCS-TC-33-050725MS	Tetrachloro-m-xylene	1	20	21.0	105		19	148
		Decachlorobiphenyl	2	20	18.4	92		20	144
		Tetrachloro-m-xylene	2	20	20.5	103		19	148
Q1984-01MSD	OU4-PCS-TC-33-050725MSD	Decachlorobiphenyl	1	20	21.2	106		20	144
		Tetrachloro-m-xylene	1	20	21.4	107		19	148
		Decachlorobiphenyl	2	20	18.9	94		20	144
		Tetrachloro-m-xylene	2	20	20.9	105		19	148
I.BLK-PD088565.D	PIBLK-PD088565.D	Decachlorobiphenyl	1	20	15.2	76		43	140
		Tetrachloro-m-xylene	1	20	19.6	98		77	126
		Decachlorobiphenyl	2	20	15.5	77		43	140
		Tetrachloro-m-xylene	2	20	18.4	92		77	126