

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/15/2025

Lab File ID: BF142393.D Time Analyzed: 15:41

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	168883	6.904	623039	8.19	317544	9.94
UPPER LIMIT	337766	7.404	1246080	8.687	635088	10.439
LOWER LIMIT	84441.5	6.404	311520	7.687	158772	9.439
EPA SAMPLE NO.						
01 TP-14	157964	6.90	599494	8.18	318296	9.93
02 TP-17	157195	6.90	594153	8.18	319877	9.93
03 TP-10	160126	6.90	587497	8.19	279138	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.