

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG No.: Q2010

Instrument ID: MSVOA_D Calibration Date/Time: 05/14/2025 12:25

Lab File ID: VD080317.D Init. Calib. Date(s): 05/06/2025 05/06/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:33 15:47

GC Column: RTX-VMS ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Dichlorodifluoromethane	50.000	49.58	0.8359985	20
Chloromethane	50.000	47.08	5.834	20
Acetone	250.000	322.96	-29.1856	20
Methylene Chloride	50.000	52.87	-5.748001	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.