

Instrument ID: **MSVOA_D**

Daily Analysis Runlog For Sequence/QC Batch ID # VD050625

Review By	Mahesh Dadoda	Review On	5/7/2025 1:30:06 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/7/2025 1:47:48 PM		
SubDirectory	VD050625	HP Acquire Method	MSVOA_D	HP Processing Method	82d050625s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133815			
Initial Calibration Stds		VP133816,VP133817,VP133818,VP133819,VP133820,VP133821			
CCC					
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133822			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VD080289.D	06 May 2025 09:57	RP/MD	Ok
2	VSTDIC005	VD080290.D	06 May 2025 13:33	RP/MD	Ok,M
3	VSTDIC010	VD080291.D	06 May 2025 13:55	RP/MD	Ok,M
4	VSTDIC020	VD080292.D	06 May 2025 14:23	RP/MD	Ok,M
5	VSTDIC050	VD080293.D	06 May 2025 14:51	RP/MD	Ok,M
6	VSTDIC100	VD080294.D	06 May 2025 15:19	RP/MD	Ok,M
7	VSTDIC150	VD080295.D	06 May 2025 15:47	RP/MD	Ok,M
8	IBLK	VD080296.D	06 May 2025 16:15	RP/MD	Ok
9	VSTDICV050	VD080297.D	06 May 2025 16:43	RP/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_D**

Daily Analysis Runlog For Sequence/QCBatch ID # VD051425

Review By	Mahesh Dadoda	Review On	5/15/2025 2:31:46 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/15/2025 2:33:28 PM		
SubDirectory	VD051425	HP Acquire Method	MSVOA_D	HP Processing Method	82d050625s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133913			
CCC		VP133914,VP133915			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VD080316.D	14 May 2025 10:44	RP/MD	Ok
2	VSTDCCC050	VD080317.D	14 May 2025 12:25	RP/MD	Ok,M
3	VD0514SBL01	VD080318.D	14 May 2025 12:59	RP/MD	Ok
4	VD0514SBS01	VD080319.D	14 May 2025 13:29	RP/MD	Ok,M
5	VD0514SBSD01	VD080320.D	14 May 2025 13:57	RP/MD	Ok,M
6	Q2004-07	VD080321.D	14 May 2025 14:38	RP/MD	Ok
7	Q2004-11	VD080322.D	14 May 2025 15:07	RP/MD	Ok
8	Q2024-01	VD080323.D	14 May 2025 15:35	RP/MD	ReRun
9	Q2022-02	VD080324.D	14 May 2025 16:03	RP/MD	ReRun
10	Q2022-04	VD080325.D	14 May 2025 16:32	RP/MD	Ok
11	Q2017-03	VD080326.D	14 May 2025 17:00	RP/MD	Ok
12	Q2017-07	VD080327.D	14 May 2025 17:28	RP/MD	Ok
13	Q2019-03	VD080328.D	14 May 2025 17:55	RP/MD	Ok
14	Q2020-03	VD080329.D	14 May 2025 18:22	RP/MD	Ok
15	Q2010-01	VD080330.D	14 May 2025 18:50	RP/MD	Ok
16	Q2010-02	VD080331.D	14 May 2025 19:17	RP/MD	Ok
17	Q2010-03	VD080332.D	14 May 2025 19:43	RP/MD	Not Ok
18	Q2010-04	VD080333.D	14 May 2025 20:10	RP/MD	Not Ok
19	Q2043-01	VD080334.D	14 May 2025 20:37	RP/MD	ReRun
20	Q2043-03	VD080335.D	14 May 2025 21:04	RP/MD	Ok,M
21	VSTDCCC050	VD080336.D	14 May 2025 21:58	RP/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Maresh Dadoda	Review On	5/16/2025 11:49:17 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM		
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133916			
Initial Calibration Stds		VP133918,VP133919,VP133920,VP133921,VP133922,VP133923			
CCC		VP133925			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133924			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022252.D	15 May 2025 07:52	SY/MD	Ok
2	VSTDIC005	VY022253.D	15 May 2025 09:46	SY/MD	Ok,M
3	VSTDIC010	VY022254.D	15 May 2025 10:16	SY/MD	Ok,M
4	VSTDIC020	VY022255.D	15 May 2025 10:39	SY/MD	Ok,M
5	VSTDIC050	VY022256.D	15 May 2025 11:02	SY/MD	Ok,M
6	VSTDIC100	VY022257.D	15 May 2025 11:24	SY/MD	Ok,M
7	VSTDIC150	VY022258.D	15 May 2025 11:47	SY/MD	Ok,M
8	VIBLK	VY022259.D	15 May 2025 12:34	SY/MD	Ok
9	VSTDICV050	VY022260.D	15 May 2025 13:20	SY/MD	Ok,M
10	VY0515SBL01	VY022261.D	15 May 2025 13:44	SY/MD	Ok
11	VY0515SBS01	VY022262.D	15 May 2025 14:07	SY/MD	Ok,M
12	VY0515SBS01	VY022263.D	15 May 2025 14:29	SY/MD	Ok,M
13	Q1984-01	VY022264.D	15 May 2025 15:12	SY/MD	Ok
14	Q1984-03	VY022265.D	15 May 2025 15:35	SY/MD	Ok
15	Q1984-05	VY022266.D	15 May 2025 15:59	SY/MD	Ok
16	Q1984-07	VY022267.D	15 May 2025 16:22	SY/MD	ReRun
17	Q1984-09	VY022268.D	15 May 2025 16:46	SY/MD	Not Ok
18	Q1984-11	VY022269.D	15 May 2025 17:09	SY/MD	Not Ok
19	Q1984-13	VY022270.D	15 May 2025 17:33	SY/MD	Not Ok
20	Q1984-15	VY022271.D	15 May 2025 17:56	SY/MD	ReRun
21	Q1982-04	VY022272.D	15 May 2025 18:19	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051525

Review By	Maresh Dadoda	Review On	5/16/2025 11:49:17 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/16/2025 11:51:38 AM		
SubDirectory	VY051525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME	STD REF.#				
Tune/Reschk	VP133916				
Initial Calibration Stds	VP133918,VP133919,VP133920,VP133921,VP133922,VP133923				
CCC	VP133925				
Internal Standard/PEM	VP131783				
ICV/I.BLK	VP133924				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q1982-05	VY022273.D	15 May 2025 18:43	SY/MD	Ok
23	Q1982-06	VY022274.D	15 May 2025 19:06	SY/MD	Ok
24	VSTDCCC050	VY022275.D	15 May 2025 19:52	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY051625

Review By	Amit Patel	Review On	5/19/2025 1:35:43 PM		
Supervise By	Maresh Dadoda	Supervise On	5/19/2025 2:39:18 PM		
SubDirectory	VY051625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133936			
CCC		VP133942,VP133943			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022276.D	16 May 2025 07:54	SY/MD	Ok
2	VSTDCCC050	VY022277.D	16 May 2025 08:24	SY/MD	Ok,M
3	VY0516SBL01	VY022278.D	16 May 2025 09:37	SY/MD	Ok
4	VY0516SBS01	VY022279.D	16 May 2025 10:09	SY/MD	Ok,M
5	VY0516SBSD01	VY022280.D	16 May 2025 10:32	SY/MD	Ok,M
6	Q1984-07RE	VY022281.D	16 May 2025 11:17	SY/MD	Confirms
7	Q1984-09	VY022282.D	16 May 2025 11:41	SY/MD	Ok
8	Q1984-11	VY022283.D	16 May 2025 12:04	SY/MD	Ok
9	Q1984-13	VY022284.D	16 May 2025 12:29	SY/MD	Ok
10	Q1984-15RE	VY022285.D	16 May 2025 12:54	SY/MD	Not Ok
11	Q1982-01	VY022286.D	16 May 2025 13:17	SY/MD	Ok
12	Q1982-02	VY022287.D	16 May 2025 13:41	SY/MD	Ok
13	Q1982-03	VY022288.D	16 May 2025 14:04	SY/MD	Ok
14	Q1982-07	VY022289.D	16 May 2025 14:28	SY/MD	Ok
15	Q1982-08	VY022290.D	16 May 2025 14:51	SY/MD	Ok
16	Q2010-03	VY022291.D	16 May 2025 15:15	SY/MD	Ok
17	Q2010-04	VY022292.D	16 May 2025 15:38	SY/MD	Ok
18	Q2003-02	VY022293.D	16 May 2025 16:02	SY/MD	ReRun
19	Q2032-03	VY022294.D	16 May 2025 16:25	SY/MD	Ok
20	Q2032-04	VY022295.D	16 May 2025 16:48	SY/MD	Ok
21	Q2032-07	VY022296.D	16 May 2025 17:12	SY/MD	Ok

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY051625

Review By	Amit Patel	Review On	5/19/2025 1:35:43 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/19/2025 2:39:18 PM		
SubDirectory	VY051625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133936			
CCC		VP133942,VP133943			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q2032-08	VY022297.D	16 May 2025 17:35	SY/MD	Ok
23	Q2032-02	VY022298.D	16 May 2025 17:59	SY/MD	Ok
24	Q2032-05MS	VY022299.D	16 May 2025 18:21	SY/MD	Not Ok
25	Q2032-06MSD	VY022300.D	16 May 2025 18:44	SY/MD	Not Ok
26	VSTDCCC050	VY022301.D	16 May 2025 19:29	SY/MD	Ok,M

M : Manual Integration