

Surrogate Summary

SDG No.: Q2010

Client: CDM Smith

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2010-01	TP-10	1,2-Dichloroethane-d4	50	69.7	139	63	155
		Dibromofluoromethane	50	60.4	121	70	134
		Toluene-d8	50	52.5	105	74	123
		4-Bromofluorobenzene	50	44.9	90	38	136
Q2010-02	TP-13	1,2-Dichloroethane-d4	50	72.5	145	63	155
		Dibromofluoromethane	50	59.4	119	70	134
		Toluene-d8	50	50.1	100	74	123
		4-Bromofluorobenzene	50	42.4	85	38	136
Q2010-03	TP-14	1,2-Dichloroethane-d4	50	56.0	112	63	155
		Dibromofluoromethane	50	52.3	105	70	134
		Toluene-d8	50	49.4	99	74	123
		4-Bromofluorobenzene	50	42.9	86	38	136
Q2010-04	TP-17	1,2-Dichloroethane-d4	50	51.1	102	63	155
		Dibromofluoromethane	50	51.0	102	70	134
		Toluene-d8	50	49.0	98	74	123
		4-Bromofluorobenzene	50	40.8	82	38	136
VD0514SBL01	VD0514SBL01	1,2-Dichloroethane-d4	50	67.2	134	63	155
		Dibromofluoromethane	50	58.6	117	70	134
		Toluene-d8	50	50.8	102	74	123
		4-Bromofluorobenzene	50	44.8	90	38	136
VD0514SBS01	VD0514SBS01	1,2-Dichloroethane-d4	50	48.4	97	63	155
		Dibromofluoromethane	50	52.2	104	70	134
		Toluene-d8	50	50.9	102	74	123
		4-Bromofluorobenzene	50	49.5	99	38	136
VD0514SBSD01	VD0514SBSD01	1,2-Dichloroethane-d4	50	47.8	96	63	155
		Dibromofluoromethane	50	52.9	106	70	134
		Toluene-d8	50	51.2	102	74	123
		4-Bromofluorobenzene	50	51.6	103	38	136
VY0516SBL01	VY0516SBL01	1,2-Dichloroethane-d4	50	43.9	88	63	155
		Dibromofluoromethane	50	48.0	96	70	134
		Toluene-d8	50	48.8	98	74	123
		4-Bromofluorobenzene	50	48.5	97	38	136
VY0516SBS01	VY0516SBS01	1,2-Dichloroethane-d4	50	51.1	102	63	155
		Dibromofluoromethane	50	50.8	102	70	134
		Toluene-d8	50	50.9	102	74	123
		4-Bromofluorobenzene	50	49.4	99	38	136