

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010
 Lab File ID: VD080317.D Date Analyzed: 05/14/2025
 Instrument ID: MSVOA_D Time Analyzed: 12:25
 GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	122311	7.88	190669	8.78	180076	11.58
UPPER LIMIT	244622	8.376	381338	9.276	360152	12.082
LOWER LIMIT	61155.5	7.376	95334.5	8.276	90038	11.082
EPA SAMPLE NO.						
TP-10	68271	7.88	175457	8.78	176652	11.58
TP-13	68027	7.88	175314	8.78	166929	11.58
VD0514SBL01	76535	7.88	201385	8.78	193499	11.58
VD0514SBS01	107481	7.88	177282	8.78	170015	11.58
VD0514SBSD01	108123	7.88	175447	8.78	170138	11.58

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.