

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q2010 SAS No.: Q2010 SDG NO.: Q2010  
 Lab File ID: VD080317.D Date Analyzed: 05/14/2025  
 Instrument ID: MSVOA\_D Time Analyzed: 12:25  
 GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 99603         | 13.517 |  |  |  |  |
| UPPER LIMIT    | 199206        | 14.017 |  |  |  |  |
| LOWER LIMIT    | 49801.5       | 13.017 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| TP-10          | 76572         | 13.52  |  |  |  |  |
| TP-13          | 69324         | 13.52  |  |  |  |  |
| VD0514SBL01    | 87202         | 13.52  |  |  |  |  |
| VD0514SBS01    | 97541         | 13.52  |  |  |  |  |
| VD0514SBSD01   | 94201         | 13.52  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.