

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL032725

Review By	SAM	Review On	4/3/2025 8:45:16 AM
Supervise By		Supervise On	
SubDirectory	VL032725	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL032725AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2589		
Initial Calibration Stds	AP2593,AP2595,AP2596		
CCC	AP2593		
Internal Standard/PEM	AP2589		
ICV/I.BLK	AP2594		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042171.D	27 Mar 2025 07:44	SY/MD	Ok
2	VSTDICCC010	VL042172.D	27 Mar 2025 08:26	SY/MD	Ok,M
3	VSTDICCC002	VL042173.D	27 Mar 2025 08:59	SY/MD	Ok,M
4	VSTDICCC001	VL042174.D	27 Mar 2025 09:30	SY/MD	Ok,M
5	VSTDICCC0.5	VL042175.D	27 Mar 2025 10:01	SY/MD	Ok,M
6	VSTDICCC0.1	VL042176.D	27 Mar 2025 10:32	SY/MD	Ok,M
7	VSTDICCC0.03	VL042177.D	27 Mar 2025 11:03	SY/MD	Ok,M
8	VSTDICCC015	VL042178.D	27 Mar 2025 11:36	SY/MD	Ok,M
9	VSTDICV010	VL042179.D	27 Mar 2025 12:23	SY/MD	Ok,M
10	VL0327ABL01	VL042180.D	27 Mar 2025 13:50	SY/MD	Ok
11	VL0327ABS01	VL042181.D	27 Mar 2025 14:39	SY/MD	Ok,M
12	QC032625 CAN 10448	VL042182.D	27 Mar 2025 16:41	SY/MD	Ok
13	Q1649-01	VL042183.D	27 Mar 2025 17:13	SY/MD	Not Ok
14	Q1649-02	VL042184.D	27 Mar 2025 17:45	SY/MD	Not Ok
15	Q1669-01	VL042185.D	27 Mar 2025 18:22	SY/MD	Dilution
16	Q1669-02	VL042186.D	27 Mar 2025 18:59	SY/MD	Ok,M
17	Q1669-03	VL042187.D	27 Mar 2025 19:36	SY/MD	Ok,M
18	Q1669-04	VL042188.D	27 Mar 2025 20:12	SY/MD	Ok,M
19	Q1669-09	VL042189.D	27 Mar 2025 20:48	SY/MD	Ok,M
20	Q1669-09DUP	VL042190.D	27 Mar 2025 21:24	SY/MD	Ok,M
21	Q1669-05	VL042191.D	27 Mar 2025 21:57	SY/MD	Not Ok

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL032725

Review By	SAM	Review On	4/3/2025 8:45:16 AM
Supervise By		Supervise On	
SubDirectory	VL032725	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL032725AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2589		
Initial Calibration Stds	AP2593,AP2595,AP2596		
CCC	AP2593		
Internal Standard/PEM	AP2589		
ICV/I.BLK	AP2594		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1669-07DL	VL042192.D	27 Mar 2025 22:29	SY/MD	Ok,M
23	Q1669-08	VL042193.D	27 Mar 2025 23:01	SY/MD	Not Ok
24	VIBLK	VL042194.D	27 Mar 2025 23:34	SY/MD	Ok
25	Q1669-01DL	VL042195.D	28 Mar 2025 00:06	SY/MD	Ok,M
26	Q1669-01DUP	VL042196.D	28 Mar 2025 00:44	SY/MD	Ok,M
27	Q1669-05	VL042197.D	28 Mar 2025 01:20	SY/MD	Ok,M
28	Q1669-07	VL042198.D	28 Mar 2025 01:57	SY/MD	Dilution
29	Q1669-08	VL042199.D	28 Mar 2025 02:33	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL041525

Review By	MMDadoda	Review On	4/16/2025 3:47:46 PM
Supervise By		Supervise On	
SubDirectory	VL041525	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL032725AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2591		
Initial Calibration Stds	AP2593,AP2595,AP2596		
CCC	AP2593		
Internal Standard/PEM	AP2591		
ICV/I.BLK	AP2594		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042325.D	15 Apr 2025 07:45	SY/MD	Ok
2	VSTDCCC010	VL042326.D	15 Apr 2025 11:30	SY/MD	Ok,M
3	VL0415ABL01	VL042327.D	15 Apr 2025 12:12	SY/MD	Ok
4	VL0415ABS01	VL042328.D	15 Apr 2025 12:54	SY/MD	Ok,M
5	QC041425 CAN 10401	VL042329.D	15 Apr 2025 13:26	SY/MD	Not Ok
6	QC041425 CAN 10401	VL042330.D	15 Apr 2025 13:57	SY/MD	Ok
7	Q1665-18DL	VL042331.D	15 Apr 2025 14:29	SY/MD	Ok
8	Q1665-10DL2	VL042332.D	15 Apr 2025 15:01	SY/MD	Ok,M
9	Q1802-02	VL042333.D	15 Apr 2025 15:35	SY/MD	Ok,M
10	Q1802-02DUP	VL042334.D	15 Apr 2025 16:08	SY/MD	Ok,M
11	Q1802-03	VL042335.D	15 Apr 2025 16:42	SY/MD	Ok,M
12	Q1802-01	VL042336.D	15 Apr 2025 17:14	SY/MD	Dilution
13	Q1802-01DL	VL042337.D	15 Apr 2025 17:46	SY/MD	Not Ok
14	Q1802-01DL	VL042338.D	15 Apr 2025 18:21	SY/MD	Ok,M

M : Manual Integration

Process Report

Order ID : Q2011



Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
05/12/2025	B2505018	Q2011-02	-5.5	10599	6 L	10525		C2504007	VL042330.D Seq :vl041525 TUNE : VL042325.D ICAL : VL042177.D CCAL : VL042326.D MB : VL042327.D
05/12/2025	B2505018	Q2011-01	-7.1	10154	6 L	10579		C2504008	VL042399.D Seq :VL042125 TUNE : VL042394.D ICAL : VL042354.D CCAL : VL042395.D MB : VL042396.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042171.D
Acq On : 27 Mar 2025 07:44
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Mar 28 02:00:52 2025



AutoFind: Scans 3177, 3178, 3179; Background Corrected with Scan 3150

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.0	87995	PASS
75	95	30	66	56.3	215296	PASS
95	95	100	100	100.0	382268	PASS
96	95	5	9	6.6	25383	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	63.8	243904	PASS
175	174	4	9	8.0	19539	PASS
176	174	93	101	95.7	233493	PASS
177	176	5	9	6.8	15877	PASS

Average of 10.370 to 10.377 min.: VL042171.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
33.15	1400	47.05	4001	57.95	183	67.00	1161
33.90	31	47.80	824	58.15	219	68.05	43530
36.15	4138	48.05	1739	58.80	27	69.10	43612
37.10	22759	49.05	18477	59.20	62	70.05	3072
38.05	19773	50.05	87995	60.05	3842	71.05	391
39.00	8558	51.05	27372	61.10	18723	71.30	148
41.00	270	52.00	1090	62.05	19083	72.05	2510
43.05	65	54.00	265	63.05	14806	73.05	17047
43.95	1610	55.05	633	64.05	1355	74.05	72175
45.00	3695	56.05	6441	64.95	108	75.10	215296
46.00	111	57.10	10669	65.95	87	76.05	17968

Average of 10.370 to 10.377 min.: VL042171.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
77.00	1756	84.70	60	96.10	25383	105.85	1266
77.90	365	85.50	69	96.95	454	106.85	241
78.00	749	85.90	91	98.20	118	108.10	78
78.90	8287	87.05	12260	98.80	119	110.00	6
79.95	2770	88.00	11240	99.15	113	110.80	163
81.00	8994	88.95	27	100.10	27	111.10	228
81.95	2105	91.05	1113	102.10	25	111.80	169
83.05	87	92.05	10732	102.80	142	112.40	56
83.30	152	93.10	15706	104.00	1429	112.60	56
83.95	3	94.10	48315	104.70	208	113.05	295
84.20	78	95.10	382268	104.95	148	114.15	45

Average of 10.370 to 10.377 min.: VL042171.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
114.90	395	123.10	107	132.40	22	141.00	3317
115.95	1131	123.85	235	132.90	67	142.00	364
116.95	1918	124.95	222	133.80	40	142.95	3384
117.90	1124	125.80	79	134.10	121	143.95	131
118.95	1429	126.10	146	134.95	750	144.60	27
119.80	193	127.10	144	135.95	199	144.95	267
120.90	128	127.95	1262	136.70	145	145.90	507
121.10	24	128.70	250	137.05	545	146.60	55
122.00	119	129.00	312	138.80	28	146.95	155
122.50	33	129.95	1135	139.10	132	147.20	55
122.90	76	130.90	590	139.95	372	148.00	938

Average of 10.370 to 10.377 min.: VL042171.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
148.80	166	156.95	650	169.10	175	179.40	63
149.05	162	158.95	474	169.90	95	191.05	70
149.95	274	160.00	21	170.15	82	203.35	48
151.00	148	160.70	162	171.10	216	208.10	24
151.80	110	160.90	233	171.55	165	219.00	22
152.25	182	161.15	346	172.30	258	236.30	19
152.95	205	161.80	102	174.00	243904	238.10	24
154.00	201	162.90	25	175.10	19539	244.05	44
155.00	271	165.50	37	176.00	233493	246.20	61
155.15	639	166.35	72	177.05	15877	250.10	21
155.90	116	167.00	175	177.90	397	270.90	18

Average of 10.370 to 10.377 min.: VL042171.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
283.40	34						
288.00	20						

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042172.D
 Acq On : 27 Mar 2025 08:26
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Quant Time: Mar 28 01:31:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	187002	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	545068	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	519536	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.371	95	422410	9.901	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.496	85	262496	9.507	ppbv	100
3) Chlorodifluoromethane	1.473	51	272522	9.662	ppbv	100
4) Chloromethane	1.531	50	107099	9.491	ppbv	100
5) Vinyl Chloride	1.577	62	102699	8.925	ppbv	100
6) Bromomethane	1.664	94	47950	9.377	ppbv	100
7) Chloroethane	1.703	64	42516	9.387	ppbv	100
8) Dichlorotetrafluoroethane	1.551	85	227041	9.656	ppbv	100
9) Propene	1.483	41	123386	9.709	ppbv	100
10) Heptane	4.972	43	335086	9.744	ppbv	100
11) Trichlorofluoromethane	1.874	101	250349	9.318	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.137	101	209365	9.481	ppbv	100
13) Ethanol	1.719	45	5338	5.364	ppbv	100
14) Bromoethene	1.777	108	77999	9.220	ppbv	100
15) Acetone	1.832	43	208061	7.955	ppbv	100
16) 1,3-Butadiene	1.606	39	107104	9.105	ppbv	100
17) tert-Butyl alcohol	2.036	59	302616	9.290	ppbv	100
18) 1,1-Dichloroethene	2.030	96	94236	9.128	ppbv	100
19) Isopropyl Alcohol	1.884	45	131359	8.902	ppbv	100
20) Methylene Chloride	2.059	84	81828	8.401	ppbv	100
21) Allyl Chloride	2.091	41	168602	9.141	ppbv	100
22) trans-1,2-Dichloroethene	2.337	96	110371	9.149	ppbv	100
23) Vinyl Acetate	2.460	43	93080	8.563	ppbv	100
24) 1,1-Dichloroethane	2.405	63	209753	9.484	ppbv	100
25) Ethyl Acetate	2.833	43	646501	9.524	ppbv	100
26) Hexane	2.823	57	249331	9.510	ppbv	100
27) Carbon Disulfide	2.146	76	268507	9.550	ppbv	100
28) Methyl tert-Butyl Ether	2.438	73	155398	9.520	ppbv	100
29) Chloroform	2.845	83	361264	9.415	ppbv	100
30) Cyclohexane	3.852	84	217215	9.537	ppbv	100
31) cis-1,2-Dichloroethene	2.716	61	256425	9.634	ppbv	100
32) 1,1,1-Trichloroethane	3.363	97	385628	8.784	ppbv	100
34) 2-Butanone	2.554	43	363735	9.203	ppbv	100
35) Carbon Tetrachloride	3.758	117	382925	9.129	ppbv	100
36) Benzene	3.651	78	508860	9.417	ppbv	100
37) 1,2-Dichloroethane	3.214	62	292937	9.469	ppbv	100
38) Trichloroethene	4.545	130	204027	9.045	ppbv	100
39) 1,2-Dichloropropane	4.312	63	183439	9.321	ppbv	100
40) 1,4-Dioxane	4.577	88	92327	9.042	ppbv #	100
41) Tetrahydrofuran	3.049	42	216851	9.643	ppbv	100
42) Bromodichloromethane	4.483	83	418887	9.822	ppbv	100
43) Methyl Methacrylate	4.872	69	219399	9.618	ppbv	100
44) 2,2,4-Trimethylpentane	4.635	57	876649	9.600	ppbv	100
45) t-1,3-Dichloropropene	6.406	75	239180	9.746	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042172.D
 Acq On : 27 Mar 2025 08:26
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Quant Time: Mar 28 01:31:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	300021	9.973	ppbv	100
47) 1,1,2-Trichloroethane	6.574	97	198176	9.526	ppbv	100
48) Dibromochloromethane	7.380	129	346499	9.828	ppbv	100
49) Bromoform	9.477	173	317189	10.021	ppbv	100
50) 4-Methyl-2-Pentanone	5.745	43	559800	8.934	ppbv	100
51) 2-Hexanone	7.432	43	460013	8.848	ppbv #	100
52) Tetrachloroethene	8.218	164	186912	8.722	ppbv	100
53) Toluene	6.927	91	613378	9.576	ppbv	100
54) 1,2-Dibromoethane	7.649	107	276797	9.536	ppbv	100
56) 1,1,1,2-Tetrachloroethane	8.917	131	265037	9.268	ppbv	100
57) Chlorobenzene	8.917	112	454511	9.223	ppbv	100
58) Ethyl Benzene	9.351	91	868003	9.406	ppbv	100
59) m/p-Xylene	9.549	91	1372806m	18.607	ppbv	
60) o-Xylene	9.960	91	689933	9.340	ppbv	100
61) Styrene	9.866	104	354697	9.706	ppbv	100
62) Isopropylbenzene	10.536	105	1019255	9.227	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.960	83	413392	9.210	ppbv	100
64) n-propylbenzene	10.986	120	259405	9.122	ppbv	100
65) tert-Butylbenzene	11.529	119	901951	8.828	ppbv	100
66) Benzyl Chloride	11.610	91	105411	9.724	ppbv	100
67) sec-Butylbenzene	11.766	105	1249106	8.796	ppbv	100
69) p-Isopropyltoluene	11.924	119	1079675	8.830	ppbv	100
70) n-Butylbenzene	12.277	91	1110351	8.879	ppbv	100
71) 2-Chlorotoluene	10.898	91	799102	9.320	ppbv	100
72) 4-Ethyltoluene	11.122	105	846232	9.200	ppbv	100
73) 1,3,5-Trimethylbenzene	11.199	105	720566	8.970	ppbv	100
74) 1,2,4-Trimethylbenzene	11.536	105	784745	8.480	ppbv	100
75) 1,3-Dichlorobenzene	11.604	146	458226	9.028	ppbv	100
76) 1,4-Dichlorobenzene	11.669	146	461371	9.134	ppbv	100
77) 1,2-Dichlorobenzene	11.944	146	441473	8.933	ppbv	100
78) Hexachloro-1,3-Butadiene	13.879	225	416735	8.284	ppbv	100
79) Naphthalene	13.510	128	938470	9.271	ppbv	100
80) Naphthalene,2-methyl-	14.455	142	514394	10.436	ppbv	100
81) 1,2,4-Trichlorobenzene	13.439	180	446420	9.561	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042172.D
Acq On : 27 Mar 2025 08:26
Operator : SY/MD
Sample : VSTDICCC010
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:31:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042173.D
 Acq On : 27 Mar 2025 08:59
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:32:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	188395	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	546394	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	504538	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 417532 10.077 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 100.800%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	55859	2.008	ppbv	99
3) Chlorodifluoromethane	1.473	51	57045	2.007	ppbv	99
4) Chloromethane	1.531	50	23293	2.049	ppbv	94
5) Vinyl Chloride	1.576	62	21989	1.897	ppbv	99
6) Bromomethane	1.664	94	10027	1.946	ppbv	98
7) Chloroethane	1.703	64	8413	1.844	ppbv	92
8) Dichlorotetrafluoroethane	1.551	85	48008	2.027	ppbv	98
9) Propene	1.483	41	26295	2.054	ppbv	94
10) Heptane	4.972	43	71466	2.063	ppbv	99
11) Trichlorofluoromethane	1.874	101	55754	2.060	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.136	101	44761	2.012	ppbv	98
13) Ethanol	1.719	45	2748m	2.741	ppbv	
14) Bromoethene	1.780	108	17323	2.033	ppbv #	89
15) Acetone	1.832	43	55365	2.101	ppbv	95
16) 1,3-Butadiene	1.606	39	22894	1.932	ppbv	95
17) tert-Butyl alcohol	2.039	59	66174	2.016	ppbv	100
18) 1,1-Dichloroethene	2.030	96	21103	2.029	ppbv	95
19) Isopropyl Alcohol	1.884	45	27804	1.870	ppbv #	35
20) Methylene Chloride	2.059	84	20518	2.091	ppbv	96
21) Allyl Chloride	2.094	41	38425	2.068	ppbv	99
22) trans-1,2-Dichloroethene	2.337	96	23384	1.924	ppbv	92
23) Vinyl Acetate	2.460	43	20736	1.894	ppbv #	97
24) 1,1-Dichloroethane	2.405	63	44740	2.008	ppbv	99
25) Ethyl Acetate	2.832	43	139516	2.040	ppbv	99
26) Hexane	2.823	57	52095	1.972	ppbv	92
27) Carbon Disulfide	2.146	76	57969	2.047	ppbv #	76
28) Methyl tert-Butyl Ether	2.437	73	33703	2.049	ppbv	99
29) Chloroform	2.845	83	78043	2.019	ppbv	95
30) Cyclohexane	3.852	84	45641	1.989	ppbv	99
31) cis-1,2-Dichloroethene	2.719	61	53713	2.003	ppbv	96
32) 1,1,1-Trichloroethane	3.363	97	81070	1.833	ppbv	99
34) 2-Butanone	2.554	43	84157	2.124	ppbv	99
35) Carbon Tetrachloride	3.758	117	81735	1.944	ppbv	96
36) Benzene	3.651	78	110727	2.044	ppbv	97
37) 1,2-Dichloroethane	3.214	62	61476	1.982	ppbv	97
38) Trichloroethene	4.541	130	43501	1.924	ppbv	97
39) 1,2-Dichloropropane	4.308	63	39912	2.023	ppbv	93
40) 1,4-Dioxane	4.587	88	20884m	2.040	ppbv	
41) Tetrahydrofuran	3.056	42	47262	2.097	ppbv	99
42) Bromodichloromethane	4.483	83	87640	2.050	ppbv	99
43) Methyl Methacrylate	4.878	69	44624	1.952	ppbv	97
44) 2,2,4-Trimethylpentane	4.632	57	188282	2.057	ppbv	100
45) t-1,3-Dichloropropene	6.406	75	49461	2.011	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042173.D
 Acq On : 27 Mar 2025 08:59
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:32:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
46) cis-1,3-Dichloropropene	5.593	75	61829	2.050	ppbv	#	91
47) 1,1,2-Trichloroethane	6.577	97	41387	1.985	ppbv		95
48) Dibromochloromethane	7.383	129	72441	2.050	ppbv		100
49) Bromoform	9.477	173	65221	2.056	ppbv		98
50) 4-Methyl-2-Pentanone	5.752	43	132328	2.107	ppbv		99
51) 2-Hexanone	7.435	43	112681	2.162	ppbv	#	99
52) Tetrachloroethene	8.221	164	40558	1.888	ppbv		98
53) Toluene	6.923	91	131314	2.045	ppbv		99
54) 1,2-Dibromoethane	7.648	107	56799	1.952	ppbv		90
56) 1,1,1,2-Tetrachloroethane	8.920	131	58141	2.093	ppbv		97
57) Chlorobenzene	8.914	112	100308	2.096	ppbv	#	94
58) Ethyl Benzene	9.351	91	191114	2.132	ppbv		96
59) m/p-Xylene	9.548	91	299531m	4.181	ppbv		
60) o-Xylene	9.959	91	151292	2.109	ppbv		98
61) Styrene	9.865	104	74134	2.089	ppbv		98
62) Isopropylbenzene	10.535	105	228191	2.127	ppbv		99
63) 1,1,2,2-Tetrachloroethane	9.956	83	90114	2.067	ppbv		99
64) n-propylbenzene	10.985	120	58105	2.104	ppbv		92
65) tert-Butylbenzene	11.529	119	216336	2.180	ppbv		99
66) Benzyl Chloride	11.610	91	22272	2.116	ppbv		95
67) sec-Butylbenzene	11.762	105	297381	2.156	ppbv		99
69) p-Isopropyltoluene	11.921	119	258357	2.176	ppbv		100
70) n-Butylbenzene	12.277	91	270616	2.228	ppbv		99
71) 2-Chlorotoluene	10.898	91	177487	2.132	ppbv		100
72) 4-Ethyltoluene	11.121	105	194660	2.179	ppbv		99
73) 1,3,5-Trimethylbenzene	11.199	105	169221	2.169	ppbv		96
74) 1,2,4-Trimethylbenzene	11.532	105	198351	2.207	ppbv		96
75) 1,3-Dichlorobenzene	11.604	146	107217	2.175	ppbv		97
76) 1,4-Dichlorobenzene	11.668	146	104925	2.139	ppbv		99
77) 1,2-Dichlorobenzene	11.943	146	102310	2.132	ppbv		99
78) Hexachloro-1,3-Butadiene	13.879	225	109539	2.242	ppbv		99
79) Naphthalene	13.510	128	233451	2.375	ppbv		97
80) Naphthalene,2-methyl-	14.455	142	117313	2.451	ppbv		99
81) 1,2,4-Trichlorobenzene	13.435	180	103765	2.288	ppbv		97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042174.D
 Acq On : 27 Mar 2025 09:30
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Quant Time: Mar 28 01:33:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	186244	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	537244	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	499830	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.371	95	404753	9.861	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	28203	1.026	ppbv	99
3) Chlorodifluoromethane	1.473	51	28693	1.021	ppbv	94
4) Chloromethane	1.531	50	11111	0.989	ppbv	99
5) Vinyl Chloride	1.577	62	10723	0.936	ppbv	98
6) Bromomethane	1.667	94	4984	0.979	ppbv	97
7) Chloroethane	1.703	64	4622	1.025	ppbv	100
8) Dichlorotetrafluoroethane	1.554	85	23886	1.020	ppbv	98
9) Propene	1.483	41	13699m	1.082	ppbv	
10) Heptane	4.975	43	36132	1.055	ppbv	100
11) Trichlorofluoromethane	1.875	101	27136	1.014	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.137	101	22629	1.029	ppbv	99
13) Ethanol	1.719	45	1610m	1.624	ppbv	
14) Bromoethene	1.781	108	8330	0.989	ppbv #	82
15) Acetone	1.836	43	29294	1.125	ppbv	95
16) 1,3-Butadiene	1.606	39	13133	1.121	ppbv	97
17) tert-Butyl alcohol	2.040	59	33500	1.033	ppbv	97
18) 1,1-Dichloroethene	2.033	96	10532	1.024	ppbv	89
19) Isopropyl Alcohol	1.884	45	14903	1.014	ppbv #	38
20) Methylene Chloride	2.059	84	10481	1.080	ppbv	93
21) Allyl Chloride	2.095	41	19039	1.036	ppbv	99
22) trans-1,2-Dichloroethene	2.341	96	13051m	1.086	ppbv	
23) Vinyl Acetate	2.464	43	12883m	1.190	ppbv	
24) 1,1-Dichloroethane	2.409	63	22505	1.022	ppbv	99
25) Ethyl Acetate	2.836	43	67402	0.997	ppbv #	96
26) Hexane	2.823	57	27223	1.043	ppbv	98
27) Carbon Disulfide	2.146	76	28167	1.006	ppbv #	47
28) Methyl tert-Butyl Ether	2.441	73	16501	1.015	ppbv	95
29) Chloroform	2.845	83	38516	1.008	ppbv	96
30) Cyclohexane	3.852	84	23052	1.016	ppbv	97
31) cis-1,2-Dichloroethene	2.716	61	27187	1.026	ppbv	89
32) 1,1,1-Trichloroethane	3.363	97	40993	0.938	ppbv	99
34) 2-Butanone	2.557	43	41423	1.063	ppbv	98
35) Carbon Tetrachloride	3.758	117	39345	0.952	ppbv	93
36) Benzene	3.651	78	55156	1.036	ppbv	94
37) 1,2-Dichloroethane	3.218	62	30543	1.002	ppbv	98
38) Trichloroethene	4.545	130	21906	0.985	ppbv	93
39) 1,2-Dichloropropane	4.312	63	20020	1.032	ppbv	99
40) 1,4-Dioxane	4.593	88	10592m	1.052	ppbv	
41) Tetrahydrofuran	3.056	42	22818	1.029	ppbv	97
42) Bromodichloromethane	4.483	83	41864	0.996	ppbv	90
43) Methyl Methacrylate	4.875	69	24233m	1.078	ppbv	
44) 2,2,4-Trimethylpentane	4.632	57	90727	1.008	ppbv	96
45) t-1,3-Dichloropropene	6.403	75	23993m	0.992	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042174.D
 Acq On : 27 Mar 2025 09:30
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:33:49 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	30063	1.014	ppbv	93
47) 1,1,2-Trichloroethane	6.577	97	20546	1.002	ppbv	96
48) Dibromochloromethane	7.383	129	34425	0.991	ppbv	100
49) Bromoform	9.481	173	31610	1.013	ppbv	98
50) 4-Methyl-2-Pentanone	5.752	43	65226m	1.056	ppbv	
51) 2-Hexanone	7.438	43	54680	1.067	ppbv #	100
52) Tetrachloroethene	8.222	164	19415	0.919	ppbv	95
53) Toluene	6.927	91	64612	1.023	ppbv	99
54) 1,2-Dibromoethane	7.652	107	28879	1.009	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.924	131	28764	1.045	ppbv #	87
57) Chlorobenzene	8.917	112	48930	1.032	ppbv	93
58) Ethyl Benzene	9.348	91	91086	1.026	ppbv	97
59) m/p-Xylene	9.529	91	147162m	2.073	ppbv	
60) o-Xylene	9.960	91	73702	1.037	ppbv	96
61) Styrene	9.866	104	36110	1.027	ppbv	97
62) Isopropylbenzene	10.536	105	109897	1.034	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.960	83	44431	1.029	ppbv	99
64) n-propylbenzene	10.986	120	29054	1.062	ppbv	98
65) tert-Butylbenzene	11.526	119	103520	1.053	ppbv	96
66) Benzyl Chloride	11.614	91	10660	1.022	ppbv #	93
67) sec-Butylbenzene	11.762	105	149660	1.095	ppbv	97
69) p-Isopropyltoluene	11.921	119	127509	1.084	ppbv	99
70) n-Butylbenzene	12.280	91	129860	1.079	ppbv	99
71) 2-Chlorotoluene	10.898	91	86356	1.047	ppbv	100
72) 4-Ethyltoluene	11.122	105	92912	1.050	ppbv	99
73) 1,3,5-Trimethylbenzene	11.199	105	83497	1.080	ppbv	92
74) 1,2,4-Trimethylbenzene	11.533	105	97858	1.099	ppbv	94
75) 1,3-Dichlorobenzene	11.604	146	51096	1.046	ppbv	99
76) 1,4-Dichlorobenzene	11.669	146	50841	1.046	ppbv	99
77) 1,2-Dichlorobenzene	11.944	146	51523	1.084	ppbv	99
78) Hexachloro-1,3-Butadiene	13.879	225	55611	1.149	ppbv	97
79) Naphthalene	13.510	128	109031	1.120	ppbv	99
80) Naphthalene,2-methyl-	14.455	142	44518	0.939	ppbv	99
81) 1,2,4-Trichlorobenzene	13.436	180	47040	1.047	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042174.D
Acq On : 27 Mar 2025 09:30
Operator : SY/MD
Sample : VSTDIC001
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC001

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:33:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042175.D
 Acq On : 27 Mar 2025 10:01
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:34:42 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	180502	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	523832	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.875	117	485078	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.370 95 403513 10.130 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	14455	0.542	ppbv	95
3) Chlorodifluoromethane	1.473	51	14637	0.538	ppbv #	89
4) Chloromethane	1.531	50	5956	0.547	ppbv	94
5) Vinyl Chloride	1.577	62	5732	0.516	ppbv	92
6) Bromomethane	1.664	94	2949	0.597	ppbv	95
7) Chloroethane	1.703	64	2582	0.591	ppbv #	87
8) Dichlorotetrafluoroethane	1.551	85	11783	0.519	ppbv	98
9) Propene	1.483	41	6622m	0.540	ppbv	
10) Heptane	4.972	43	16673	0.502	ppbv	89
11) Trichlorofluoromethane	1.874	101	14099	0.544	ppbv	92
12) 1,1,2-Trichlorotrifluo...	2.136	101	11252	0.528	ppbv	99
13) Ethanol	1.722	45	997	1.038	ppbv #	36
14) Bromoethene	1.780	108	4770	0.584	ppbv #	86
15) Acetone	1.835	43	15331	0.607	ppbv #	85
16) 1,3-Butadiene	1.606	39	8594	0.757	ppbv	82
17) tert-Butyl alcohol	2.039	59	17467	0.556	ppbv	97
18) 1,1-Dichloroethene	2.030	96	5634	0.565	ppbv #	92
19) Isopropyl Alcohol	1.884	45	9171	0.644	ppbv #	39
20) Methylene Chloride	2.059	84	5634	0.599	ppbv #	93
21) Allyl Chloride	2.091	41	9632	0.541	ppbv	89
22) trans-1,2-Dichloroethene	2.334	96	6353	0.546	ppbv	83
23) Vinyl Acetate	2.457	43	6809	0.649	ppbv #	91
24) 1,1-Dichloroethane	2.405	63	11258	0.527	ppbv	97
25) Ethyl Acetate	2.836	43	36258	0.553	ppbv	98
26) Hexane	2.823	57	13605	0.538	ppbv	92
27) Carbon Disulfide	2.146	76	14510	0.535	ppbv #	1
28) Methyl tert-Butyl Ether	2.437	73	8310	0.527	ppbv #	89
29) Chloroform	2.845	83	20149	0.544	ppbv	98
30) Cyclohexane	3.852	84	12063m	0.549	ppbv	
31) cis-1,2-Dichloroethene	2.719	61	13601	0.529	ppbv	92
32) 1,1,1-Trichloroethane	3.360	97	20567	0.485	ppbv	100
34) 2-Butanone	2.557	43	19739	0.520	ppbv	96
35) Carbon Tetrachloride	3.755	117	18558	0.460	ppbv	93
36) Benzene	3.651	78	27488	0.529	ppbv #	87
37) 1,2-Dichloroethane	3.214	62	16062	0.540	ppbv	100
38) Trichloroethene	4.538	130	11460m	0.529	ppbv	
39) 1,2-Dichloropropane	4.308	63	10123	0.535	ppbv	92
40) 1,4-Dioxane	4.593	88	6121m	0.624	ppbv	
41) Tetrahydrofuran	3.059	42	11509m	0.533	ppbv	
42) Bromodichloromethane	4.483	83	20374	0.497	ppbv #	98
43) Methyl Methacrylate	4.875	69	11583m	0.528	ppbv	
44) 2,2,4-Trimethylpentane	4.632	57	46646	0.532	ppbv	95
45) t-1,3-Dichloropropene	6.406	75	12404	0.526	ppbv #	79

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042175.D
 Acq On : 27 Mar 2025 10:01
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
 Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:34:42 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	14374m	0.497	ppbv	
47) 1,1,2-Trichloroethane	6.574	97	10972m	0.549	ppbv	
48) Dibromochloromethane	7.380	129	17042	0.503	ppbv	99
49) Bromoform	9.480	173	14828	0.487	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	33674	0.559	ppbv	97
51) 2-Hexanone	7.438	43	26154	0.523	ppbv #	100
52) Tetrachloroethene	8.218	164	10502	0.510	ppbv	94
53) Toluene	6.923	91	32444	0.527	ppbv	98
54) 1,2-Dibromoethane	7.648	107	14220	0.510	ppbv	99
56) 1,1,1,2-Tetrachloroethane	8.920	131	13959	0.523	ppbv #	1
57) Chlorobenzene	8.914	112	24826	0.540	ppbv #	91
58) Ethyl Benzene	9.348	91	44862	0.521	ppbv	96
59) m/p-Xylene	9.542	91	73028m	1.060	ppbv	
60) o-Xylene	9.959	91	36879	0.535	ppbv	98
61) Styrene	9.866	104	17198	0.504	ppbv	96
62) Isopropylbenzene	10.536	105	55531	0.538	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.959	83	21989	0.525	ppbv	96
64) n-propylbenzene	10.985	120	14030	0.528	ppbv	95
65) tert-Butylbenzene	11.526	119	53115	0.557	ppbv	100
66) Benzyl Chloride	11.613	91	5263	0.520	ppbv	97
67) sec-Butylbenzene	11.762	105	72879	0.550	ppbv	98
69) p-Isopropyltoluene	11.921	119	61933	0.542	ppbv	100
70) n-Butylbenzene	12.277	91	61507	0.527	ppbv	99
71) 2-Chlorotoluene	10.895	91	41637	0.520	ppbv	95
72) 4-Ethyltoluene	11.121	105	44473	0.518	ppbv #	94
73) 1,3,5-Trimethylbenzene	11.199	105	39190	0.523	ppbv	97
74) 1,2,4-Trimethylbenzene	11.532	105	48370	0.560	ppbv	97
75) 1,3-Dichlorobenzene	11.600	146	25122	0.530	ppbv	100
76) 1,4-Dichlorobenzene	11.668	146	24991	0.530	ppbv	97
77) 1,2-Dichlorobenzene	11.943	146	24548	0.532	ppbv	96
78) Hexachloro-1,3-Butadiene	13.879	225	25636	0.546	ppbv	99
79) Naphthalene	13.510	128	47990	0.508	ppbv	95
80) Naphthalene,2-methyl-	14.458	142	17486	0.380	ppbv	98
81) 1,2,4-Trichlorobenzene	13.436	180	19924	0.457	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042175.D
Acq On : 27 Mar 2025 10:01
Operator : SY/MD
Sample : VSTDICC0.5
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:34:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042176.D
 Acq On : 27 Mar 2025 10:32
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
 Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:35:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	180740	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	528028	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	481506	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	398069	10.067	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.700%
Target Compounds						
5) Vinyl Chloride	1.580	62	1274	0.115	ppbv #	42
32) 1,1,1-Trichloroethane	3.366	97	4509m	0.106	ppbv	
35) Carbon Tetrachloride	3.768	117	4131m	0.102	ppbv	
38) Trichloroethene	4.551	130	2306m	0.106	ppbv	
52) Tetrachloroethene	8.225	164	1895m	0.091	ppbv	
54) 1,2-Dibromoethane	7.652	107	3033m	0.108	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	3852m	0.093	ppbv	
79) Naphthalene	13.513	128	7882	0.084	ppbv #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042176.D
Acq On : 27 Mar 2025 10:32
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:35:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042177.D
 Acq On : 27 Mar 2025 11:03
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:36:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	178058	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.955	114	512915	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	475617	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	389935	9.983	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	471m	0.043	ppbv	
32) 1,1,1-Trichloroethane	3.366	97	1683m	0.040	ppbv	
35) Carbon Tetrachloride	3.758	117	1645m	0.042	ppbv	
38) Trichloroethene	4.541	130	712m	0.034	ppbv	
52) Tetrachloroethene	8.208	164	880m	0.044	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.963	83	1407m	0.034	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042177.D
Acq On : 27 Mar 2025 11:03
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Mar 28 01:36:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
Supervised By : Mahesh Dadoda 04/03/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042178.D
 Acq On : 27 Mar 2025 11:36
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Quant Time: Mar 28 01:37:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Mar 28 01:31:23 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	173372	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.952	114	499360	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.878	117	473077	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	387777	9.981	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.800%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.495	85	358962	14.023	ppbv	99
3) Chlorodifluoromethane	1.473	51	366164	14.002	ppbv	98
4) Chloromethane	1.531	50	148133	14.160	ppbv	98
5) Vinyl Chloride	1.576	62	141938	13.304	ppbv	97
6) Bromomethane	1.667	94	65113	13.734	ppbv	99
7) Chloroethane	1.703	64	58820	14.007	ppbv	99
8) Dichlorotetrafluoroethane	1.550	85	314834	14.442	ppbv	98
9) Propene	1.482	41	163799	13.902	ppbv	95
10) Heptane	4.972	43	446949	14.019	ppbv	99
11) Trichlorofluoromethane	1.874	101	349967	14.050	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.136	101	295217	14.419	ppbv	99
13) Ethanol	1.719	45	7382	8.001	ppbv	92
14) Bromoethene	1.780	108	106436	13.571	ppbv	98
15) Acetone	1.832	43	296254	12.218	ppbv	98
16) 1,3-Butadiene	1.605	39	150122	13.766	ppbv	100
17) tert-Butyl alcohol	2.039	59	416362	13.787	ppbv	99
18) 1,1-Dichloroethene	2.029	96	131720	13.762	ppbv	99
19) Isopropyl Alcohol	1.884	45	179117	13.093	ppbv	100
20) Methylene Chloride	2.059	84	113183	12.533	ppbv	95
21) Allyl Chloride	2.094	41	239440	14.002	ppbv	99
22) trans-1,2-Dichloroethene	2.337	96	149455	13.362	ppbv	98
23) Vinyl Acetate	2.460	43	138419	13.736	ppbv #	96
24) 1,1-Dichloroethane	2.405	63	298760	14.570	ppbv	99
25) Ethyl Acetate	2.832	43	871882	13.855	ppbv	100
26) Hexane	2.822	57	344636	14.178	ppbv	98
27) Carbon Disulfide	2.146	76	370022	14.196	ppbv	97
28) Methyl tert-Butyl Ether	2.437	73	216463	14.303	ppbv	95
29) Chloroform	2.845	83	508614	14.297	ppbv	100
30) Cyclohexane	3.852	84	296458	14.040	ppbv	99
31) cis-1,2-Dichloroethene	2.719	61	351945	14.262	ppbv	97
32) 1,1,1-Trichloroethane	3.363	97	547329	13.448	ppbv	99
34) 2-Butanone	2.554	43	497020	13.726	ppbv	99
35) Carbon Tetrachloride	3.758	117	543385	14.140	ppbv	100
36) Benzene	3.651	78	699549	14.131	ppbv	98
37) 1,2-Dichloroethane	3.214	62	416544	14.697	ppbv	99
38) Trichloroethene	4.541	130	283430	13.715	ppbv	98
39) 1,2-Dichloropropane	4.308	63	257891	14.304	ppbv	99
40) 1,4-Dioxane	4.577	88	124925	13.354	ppbv #	91
41) Tetrahydrofuran	3.049	42	294634	14.301	ppbv	100
42) Bromodichloromethane	4.483	83	587772	15.043	ppbv	99
43) Methyl Methacrylate	4.874	69	295300	14.130	ppbv	99
44) 2,2,4-Trimethylpentane	4.635	57	1180275	14.108	ppbv	99
45) t-1,3-Dichloropropene	6.405	75	329360	14.649	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042178.D
 Acq On : 27 Mar 2025 11:36
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/31/2025
 Supervised By : Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:37:01 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 01:31:23 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.593	75	400813	14.542	ppbv	98
47) 1,1,2-Trichloroethane	6.577	97	273137	14.331	ppbv	98
48) Dibromochloromethane	7.383	129	482431	14.936	ppbv	100
49) Bromoform	9.480	173	427143	14.730	ppbv	99
50) 4-Methyl-2-Pentanone	5.745	43	755146	13.155	ppbv	99
51) 2-Hexanone	7.431	43	619419	13.005	ppbv #	99
52) Tetrachloroethene	8.221	164	258356	13.159	ppbv	95
53) Toluene	6.926	91	829648	14.137	ppbv	100
54) 1,2-Dibromoethane	7.645	107	381823	14.359	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.920	131	365410	14.032	ppbv	99
57) Chlorobenzene	8.917	112	618201	13.777	ppbv	100
58) Ethyl Benzene	9.351	91	1167096	13.889	ppbv	97
59) m/p-Xylene	9.548	91	1857561m	27.651	ppbv	
60) o-Xylene	9.962	91	913234	13.577	ppbv	99
61) Styrene	9.865	104	474195	14.250	ppbv	99
62) Isopropylbenzene	10.535	105	1362416	13.544	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.959	83	551386	13.491	ppbv	99
64) n-propylbenzene	10.985	120	356118	13.753	ppbv	94
65) tert-Butylbenzene	11.529	119	1200149	12.901	ppbv	98
66) Benzyl Chloride	11.613	91	134397	13.615	ppbv	100
67) sec-Butylbenzene	11.765	105	1643629	12.711	ppbv	99
69) p-Isopropyltoluene	11.924	119	1437016	12.906	ppbv	98
70) n-Butylbenzene	12.280	91	1477443	12.975	ppbv	99
71) 2-Chlorotoluene	10.898	91	1071432	13.724	ppbv	99
72) 4-Ethyltoluene	11.121	105	1136415	13.569	ppbv	99
73) 1,3,5-Trimethylbenzene	11.202	105	979783	13.395	ppbv	99
74) 1,2,4-Trimethylbenzene	11.535	105	1048874	12.447	ppbv	92
75) 1,3-Dichlorobenzene	11.607	146	626015	13.545	ppbv	99
76) 1,4-Dichlorobenzene	11.668	146	628694	13.668	ppbv	100
77) 1,2-Dichlorobenzene	11.947	146	602844	13.396	ppbv	100
78) Hexachloro-1,3-Butadiene	13.882	225	556392	12.147	ppbv	99
79) Naphthalene	13.513	128	1258479	13.653	ppbv	98
80) Naphthalene,2-methyl-	14.455	142	718900	16.018	ppbv	98
81) 1,2,4-Trichlorobenzene	13.439	180	598416	14.075	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
Data File : VL042178.D
Acq On : 27 Mar 2025 11:36
Operator : SY/MD
Sample : VSTDIC015
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC015

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/31/2025
Supervised By :Mahesh Dadoda 04/03/2025

Quant Time: Mar 28 01:37:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Mar 28 01:31:23 2025
Response via : Initial Calibration





Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
Data File : VL042325.D
Acq On : 15 Apr 2025 07:45
Operator : SY/MD
Sample : BFB
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
Last Update : Fri Mar 28 02:00:52 2025



AutoFind: Scans 3185, 3186, 3187; Background Corrected with Scan 3171

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.0	62363	PASS
75	95	30	66	55.5	144299	PASS
95	95	100	100	100.0	260117	PASS
96	95	5	9	6.7	17331	PASS
173	174	0.00	2	1.3	2038	PASS
174	95	50	120	62.4	162261	PASS
175	174	4	9	6.8	11092	PASS
176	174	93	101	94.8	153792	PASS
177	176	5	9	6.2	9603	PASS

Average of 10.396 to 10.403 min.: VL042325.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
33.10	527	45.05	2755	53.90	269	64.05	135
33.80	64	45.90	103	54.90	810	64.95	185
33.90	109	46.15	212	56.00	4390	65.80	2
36.00	2597	47.05	2784	57.05	7934	66.10	17
37.00	16567	48.05	1901	57.90	208	66.90	705
38.05	14005	49.05	12367	58.10	142	67.20	281
39.00	6748	50.00	62363	59.10	80	68.05	29680
41.00	129	51.05	19528	60.00	2756	69.05	29389
42.05	149	52.05	1034	61.00	13383	70.05	2291
43.00	410	52.80	160	62.00	13531	71.10	181
44.05	1955	53.10	160	63.05	12163	72.00	1472

Average of 10.396 to 10.403 min.: VL042325.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
73.05	11788	83.20	99	95.10	260117	109.95	85
74.05	46619	85.00	21	96.05	17331	110.90	165
75.05	144299	85.70	93	97.05	488	111.95	308
76.05	11834	86.15	253	103.10	144	113.00	56
77.00	1527	86.95	8157	103.90	952	113.15	76
77.95	1078	87.95	8281	104.90	256	114.70	76
78.95	7127	89.70	79	105.90	1018	114.80	86
80.00	2469	91.05	904	106.60	86	114.95	162
80.95	7534	92.00	7175	107.15	202	115.90	1065
81.90	1563	93.05	12057	107.95	121	116.95	1597
82.90	327	94.05	29936	109.70	54	117.90	910

Average of 10.396 to 10.403 min.: VL042325.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
118.95	1420	130.80	65	142.20	102	152.85	224
121.95	46	131.05	349	143.00	2519	153.85	232
123.95	170	133.80	30	144.00	295	154.90	180
124.95	99	134.80	291	145.00	595	155.10	426
125.90	212	135.10	198	145.95	414	156.00	154
126.50	20	136.70	159	146.60	18	156.85	198
126.80	63	137.05	298	147.10	125	157.10	194
127.80	419	138.75	100	147.85	450	157.95	188
128.10	269	139.80	51	148.95	106	158.90	274
128.95	251	140.95	2305	149.95	353	160.30	44
129.95	810	141.85	362	151.90	26	160.80	239

Average of 10.396 to 10.403 min.: VL042325.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
170.25	244	194.60	42				
170.80	64	209.60	24				
171.10	110						
172.00	2175						
173.05	2038						
174.00	162261						
175.00	11092						
176.00	153792						
177.00	9603						
177.90	283						
180.90	23						

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/16/2025

Quant Time: Apr 16 01:29:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.806	49	172263	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	3.988	114	457227	10.000	ppbv	0.04
55) Chlorobenzene-d5	8.911	117	411493	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.403 95 325960 9.646 ppbv 0.03
 Spiked Amount 10.000 Range 65 - 135 Recovery = 96.500%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.505	85	214541	8.435	ppbv	99
3) Chlorodifluoromethane	1.483	51	235250	9.054	ppbv	98
4) Chloromethane	1.541	50	89653	8.625	ppbv	98
5) Vinyl Chloride	1.589	62	83204	7.554	ppbv	96
6) Bromomethane	1.677	94	36297	7.705	ppbv	99
7) Chloroethane	1.716	64	32145	7.704	ppbv	97
8) Dichlorotetrafluoroethane	1.564	85	182884	8.443	ppbv	94
9) Propene	1.492	41	101034	8.483	ppbv	98
10) Heptane	5.020	43	276748	8.736	ppbv	98
11) Trichlorofluoromethane	1.887	101	202387	8.178	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.153	101	157030	7.719	ppbv	98
13) Ethanol	1.732	45	4440	3.944	ppbv	93
14) Bromoethene	1.793	108	59588	7.646	ppbv	99
15) Acetone	1.845	43	183439	7.614	ppbv	96
16) 1,3-Butadiene	1.619	39	90954	7.731	ppbv	100
17) tert-Butyl alcohol	2.052	59	209737	6.990	ppbv	100
18) 1,1-Dichloroethene	2.046	96	72322	7.605	ppbv	96
19) Isopropyl Alcohol	1.897	45	104424	7.682	ppbv	98
20) Methylene Chloride	2.075	84	60337	6.725	ppbv	90
21) Allyl Chloride	2.111	41	132956	7.825	ppbv	98
22) trans-1,2-Dichloroethene	2.357	96	81572	7.422	ppbv	96
23) Vinyl Acetate	2.480	43	72160	6.920	ppbv	97
24) 1,1-Dichloroethane	2.425	63	169411	8.315	ppbv	100
25) Ethyl Acetate	2.855	43	564611	9.030	ppbv	99
26) Hexane	2.845	57	213782	8.851	ppbv	98
27) Carbon Disulfide	2.166	76	195452	7.547	ppbv	97
28) Methyl tert-Butyl Ether	2.457	73	116301	7.734	ppbv	96
29) Chloroform	2.868	83	314164	8.888	ppbv	100
30) Cyclohexane	3.884	84	173360	8.266	ppbv	98
31) cis-1,2-Dichloroethene	2.739	61	208160	8.490	ppbv	96
32) 1,1,1-Trichloroethane	3.392	97	311758	7.704	ppbv	98
34) 2-Butanone	2.577	43	317117	9.565	ppbv	97
35) Carbon Tetrachloride	3.790	117	317942	8.902	ppbv	99
36) Benzene	3.684	78	424254	9.360	ppbv	94
37) 1,2-Dichloroethane	3.243	62	239556	9.231	ppbv	100
38) Trichloroethene	4.583	130	172146	9.102	ppbv	96
39) 1,2-Dichloropropane	4.350	63	151707	9.190	ppbv	96
40) 1,4-Dioxane	4.622	88	72075	8.226	ppbv #	98
41) Tetrahydrofuran	3.075	42	181471	9.505	ppbv	95
42) Bromodichloromethane	4.525	83	333167	9.313	ppbv	97
43) Methyl Methacrylate	4.923	69	155399	8.098	ppbv	96
44) 2,2,4-Trimethylpentane	4.677	57	728571	9.511	ppbv	99
45) t-1,3-Dichloropropene	6.461	75	151063	7.338	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/16/2025
 Supervised By :Mahesh Dadoda 04/16/2025

Quant Time: Apr 16 01:29:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.648	75	204339	8.097	ppbv	98
47) 1,1,2-Trichloroethane	6.626	97	163132	9.348	ppbv	97
48) Dibromochloromethane	7.425	129	272592	9.217	ppbv	99
49) Bromoform	9.513	173	238880	8.997	ppbv	100
50) 4-Methyl-2-Pentanone	5.800	43	442851	8.428	ppbv	97
51) 2-Hexanone	7.474	43	353471	8.192	ppbv #	97
52) Tetrachloroethene	8.260	164	144481	8.037	ppbv	98
53) Toluene	6.975	91	461854	8.595	ppbv	100
54) 1,2-Dibromoethane	7.687	107	220075	9.047	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.956	131	214175	9.455	ppbv	99
57) Chlorobenzene	8.953	112	375357	9.617	ppbv	97
58) Ethyl Benzene	9.383	91	639857	8.754	ppbv	99
59) m/p-Xylene	9.581	91	1063509m	18.222	ppbv	
60) o-Xylene	9.992	91	528896	9.040	ppbv	95
61) Styrene	9.898	104	245300	8.474	ppbv	97
62) Isopropylbenzene	10.565	105	794408	9.079	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.992	83	343786	9.671	ppbv	100
64) n-propylbenzene	11.015	120	212879	9.452	ppbv	95
65) tert-Butylbenzene	11.555	119	721078	8.911	ppbv	97
66) Benzyl Chloride	11.639	91	54427	6.339	ppbv	98
67) sec-Butylbenzene	11.791	105	1026079	9.123	ppbv	99
69) p-Isopropyltoluene	11.947	119	865385	8.935	ppbv	99
70) n-Butylbenzene	12.303	91	884265	8.928	ppbv	99
71) 2-Chlorotoluene	10.927	91	599592	8.830	ppbv	98
72) 4-Ethyltoluene	11.147	105	657765	9.029	ppbv	99
73) 1,3,5-Trimethylbenzene	11.228	105	557593	8.764	ppbv	99
74) 1,2,4-Trimethylbenzene	11.558	105	623563	8.507	ppbv #	86
75) 1,3-Dichlorobenzene	11.630	146	381572	9.492	ppbv	98
76) 1,4-Dichlorobenzene	11.694	146	380035	9.499	ppbv	98
77) 1,2-Dichlorobenzene	11.969	146	379896	9.706	ppbv	97
78) Hexachloro-1,3-Butadiene	13.902	225	311900	7.828	ppbv	99
79) Naphthalene	13.533	128	760876	9.490	ppbv	99
80) Naphthalene,2-methyl-	14.478	142	365398	9.294	ppbv	98
81) 1,2,4-Trichlorobenzene	13.461	180	337032	9.114	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042327.D
 Acq On : 15 Apr 2025 12:12
 Operator : SY/MD
 Sample : VL0415ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0415ABL01

Quant Time: Apr 16 01:30:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.813	49	173729	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	3.994	114	447320	10.000	ppbv	0.04
55) Chlorobenzene-d5	8.917	117	385318	10.000	ppbv	0.04

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.406	95	292759	9.252	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.500%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
Data File : VL042327.D
Acq On : 15 Apr 2025 12:12
Operator : SY/MD
Sample : VL0415ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0415ABL01

Quant Time: Apr 16 01:30:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042328.D
 Acq On : 15 Apr 2025 12:54
 Operator : SY/MD
 Sample : VL0415ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0415ABS01

Quant Time: Apr 16 01:30:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/16/2025
 Supervised By : Mahesh Dadoda 04/16/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.810	49	171986	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	3.994	114	444595	10.000	ppbv	0.04
55) Chlorobenzene-d5	8.917	117	399604	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.406	95	318902	9.718	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.509	85	214241	8.437	ppbv	99
3) Chlorodifluoromethane	1.486	51	234966	9.058	ppbv	97
4) Chloromethane	1.541	50	88230	8.502	ppbv	98
5) Vinyl Chloride	1.589	62	82221	7.476	ppbv	94
6) Bromomethane	1.680	94	36050	7.665	ppbv	100
7) Chloroethane	1.716	64	33039	7.931	ppbv	97
8) Dichlorotetrafluoroethane	1.564	85	178173	8.239	ppbv	95
9) Propene	1.492	41	94136	7.917	ppbv	99
10) Heptane	5.027	43	267288	8.451	ppbv	99
11) Trichlorofluoromethane	1.890	101	201665	8.162	ppbv	99
12) 1,1,2-Trichlorotrifluo...	2.156	101	154306	7.598	ppbv	98
13) Ethanol	1.732	45	4407	3.921	ppbv	97
14) Bromoethene	1.793	108	57635	7.408	ppbv	99
15) Acetone	1.848	43	178724	7.430	ppbv	96
16) 1,3-Butadiene	1.619	39	89479	7.617	ppbv	96
17) tert-Butyl alcohol	2.055	59	202421	6.757	ppbv	99
18) 1,1-Dichloroethene	2.049	96	69013	7.269	ppbv	97
19) Isopropyl Alcohol	1.900	45	100289	7.390	ppbv	97
20) Methylene Chloride	2.075	84	60054	6.704	ppbv	88
21) Allyl Chloride	2.111	41	127800	7.534	ppbv	98
22) trans-1,2-Dichloroethene	2.360	96	78991	7.198	ppbv	98
23) Vinyl Acetate	2.483	43	64777	6.222	ppbv	98
24) 1,1-Dichloroethane	2.428	63	168373	8.277	ppbv	99
25) Ethyl Acetate	2.861	43	545140	8.732	ppbv	99
26) Hexane	2.848	57	205629	8.528	ppbv	97
27) Carbon Disulfide	2.166	76	191388	7.402	ppbv	99
28) Methyl tert-Butyl Ether	2.460	73	107667	7.172	ppbv	96
29) Chloroform	2.874	83	309262	8.763	ppbv	100
30) Cyclohexane	3.891	84	165914	7.924	ppbv	98
31) cis-1,2-Dichloroethene	2.742	61	197651	8.074	ppbv	99
32) 1,1,1-Trichloroethane	3.395	97	304143	7.528	ppbv	97
34) 2-Butanone	2.580	43	300576	9.323	ppbv	98
35) Carbon Tetrachloride	3.797	117	314849	9.065	ppbv	99
36) Benzene	3.690	78	407959	9.256	ppbv	98
37) 1,2-Dichloroethane	3.247	62	235609	9.337	ppbv	99
38) Trichloroethene	4.590	130	164664	8.954	ppbv	96
39) 1,2-Dichloropropane	4.354	63	149260	9.299	ppbv	97
40) 1,4-Dioxane	4.625	88	68773	8.072	ppbv #	95
41) Tetrahydrofuran	3.082	42	173776	9.361	ppbv	95
42) Bromodichloromethane	4.532	83	327497	9.414	ppbv	99
43) Methyl Methacrylate	4.930	69	151531	8.121	ppbv	98
44) 2,2,4-Trimethylpentane	4.680	57	707832	9.503	ppbv	98
45) t-1,3-Dichloropropene	6.470	75	143714	7.179	ppbv	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042328.D
 Acq On : 15 Apr 2025 12:54
 Operator : SY/MD
 Sample : VL0415ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0415ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/16/2025
 Supervised By :Mahesh Dadoda 04/16/2025

Quant Time: Apr 16 01:30:27 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Mar 28 02:00:52 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.658	75	192419	7.841	ppbv	96
47) 1,1,2-Trichloroethane	6.635	97	163251	9.621	ppbv	98
48) Dibromochloromethane	7.428	129	266192	9.256	ppbv	100
49) Bromoform	9.519	173	236686	9.168	ppbv	99
50) 4-Methyl-2-Pentanone	5.807	43	429988	8.416	ppbv	95
51) 2-Hexanone	7.480	43	344172	8.203	ppbv #	96
52) Tetrachloroethene	8.263	164	139845	8.000	ppbv	95
53) Toluene	6.978	91	447178	8.559	ppbv	99
54) 1,2-Dibromoethane	7.694	107	214982	9.089	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.959	131	216770	9.855	ppbv	98
57) Chlorobenzene	8.956	112	369132	9.739	ppbv	95
58) Ethyl Benzene	9.390	91	618669	8.716	ppbv	100
59) m/p-Xylene	9.584	91	1049238m	18.512	ppbv	
60) o-Xylene	9.995	91	517471	9.107	ppbv	96
61) Styrene	9.904	104	229601	8.168	ppbv	97
62) Isopropylbenzene	10.568	105	785673	9.247	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.995	83	351308	10.176	ppbv	99
64) n-propylbenzene	11.018	120	210294	9.615	ppbv	98
65) tert-Butylbenzene	11.558	119	718948	9.149	ppbv	97
66) Benzyl Chloride	11.642	91	55150	6.614	ppbv	99
67) sec-Butylbenzene	11.795	105	1018861	9.328	ppbv	100
69) p-Isopropyltoluene	11.950	119	863264	9.179	ppbv	99
70) n-Butylbenzene	12.306	91	891883	9.273	ppbv	99
71) 2-Chlorotoluene	10.930	91	590408	8.953	ppbv	99
72) 4-Ethyltoluene	11.150	105	652962	9.230	ppbv	98
73) 1,3,5-Trimethylbenzene	11.231	105	559451	9.055	ppbv	100
74) 1,2,4-Trimethylbenzene	11.565	105	632271	8.883	ppbv	97
75) 1,3-Dichlorobenzene	11.633	146	381882	9.782	ppbv	98
76) 1,4-Dichlorobenzene	11.697	146	383116	9.861	ppbv	98
77) 1,2-Dichlorobenzene	11.973	146	386768	10.175	ppbv	98
78) Hexachloro-1,3-Butadiene	13.905	225	314860	8.138	ppbv	100
79) Naphthalene	13.539	128	777722	9.989	ppbv	99
80) Naphthalene,2-methyl-	14.481	142	375541	9.836	ppbv	98
81) 1,2,4-Trichlorobenzene	13.465	180	343011	9.551	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042330.D
 Acq On : 15 Apr 2025 13:57
 Operator : SY/MD
 Sample : QC041425 CAN 10401
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC041425 CAN 10401

Quant Time: Apr 16 01:31:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.813	49	162884	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	3.997	114	405181	10.000	ppbv	0.05
55) Chlorobenzene-d5	8.917	117	353699	10.000	ppbv	0.04

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.406	95	273327	9.410	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
Data File : VL042330.D
Acq On : 15 Apr 2025 13:57
Operator : SY/MD
Sample : QC041425 CAN 10401
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041425 CAN 10401

Quant Time: Apr 16 01:31:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Mar 28 02:00:52 2025
Response via : Initial Calibration

