

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q2011 SAS No.: Q2011 SDG No.: Q2011

Instrument ID: MSVOA_L Calibration Date/Time: 05/14/2025 10:23

Lab File ID: VL042536.D Init. Calib. Date(s): 05/12/2025 05/12/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:57 14:28

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.678	0.550		-18.88	30
1,1-Dichloroethane	1.244	1.181		-5.06	30
cis-1,2-Dichloroethene	1.301	1.220		-6.23	30
1,1,1-Trichloroethane	2.465	2.166		-12.13	30
Trichloroethene	0.385	0.350		-9.09	30
Tetrachloroethene	0.339	0.304		-10.32	30
1-Bromo-4-Fluorobenzene	0.870	0.871		0.12	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.