

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL051225

Review By	Semsettin Yesilyurt	Review On	5/14/2025 9:07:47 AM
Supervise By	Maresh Dadoda	Supervise On	5/14/2025 12:32:58 PM
SubDirectory	VL051225	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL051225AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2604
Initial Calibration Stds	AP2610,AP2611,AP2612
CCC	
Internal Standard/PEM	AP2604
ICV/I.BLK	AP2613
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042524.D	12 May 2025 07:58		SY/MD	Ok
2	VSTDICCC010	VSTDICCC010	VL042525.D	12 May 2025 08:57		SY/MD	Ok,M
3	VSTDICC002	VSTDICC002	VL042526.D	12 May 2025 09:33	Not used	SY/MD	Not Ok
4	VSTDICC001	VSTDICC001	VL042527.D	12 May 2025 10:07	Method fail for comp.#9,61,79,80	SY/MD	Ok,M
5	VSTDICC0.5	VSTDICC0.5	VL042528.D	12 May 2025 10:41		SY/MD	Ok,M
6	VSTDICC0.1	VSTDICC0.1	VL042529.D	12 May 2025 11:15		SY/MD	Ok,M
7	VSTDICC0.03	VSTDICC0.03	VL042530.D	12 May 2025 11:49		SY/MD	Ok,M
8	VSTDICCC010	VSTDICCC010	VL042531.D	12 May 2025 12:22	Not used	SY/MD	Not Ok
9	VSTDICC002	VSTDICC002	VL042532.D	12 May 2025 12:58		SY/MD	Ok,M
10	VSTDICC015	VSTDICC015	VL042533.D	12 May 2025 14:28		SY/MD	Ok,M
11	VSTDICV010	ICVVL051225	VL042534.D	12 May 2025 16:10	ICV Failed for comp. #51,61,66,79,80,81	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL051425

Review By	Semsettin Yesilyurt	Review On	5/15/2025 10:12:52 AM
Supervise By	Maresh Dadoda	Supervise On	5/15/2025 2:31:22 PM
SubDirectory	VL051425	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL051225AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2614
Initial Calibration Stds	AP2610,AP2611,AP2612
CCC	AP2616
Internal Standard/PEM	AP2614
ICV/I.BLK	AP2613
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042535.D	14 May 2025 09:38		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL042536.D	14 May 2025 10:23		SY/MD	Ok,M
3	VL0514ABL01	VL0514ABL01	VL042537.D	14 May 2025 11:09		SY/MD	Ok
4	VL0514ABS01	VL0514ABS01	VL042538.D	14 May 2025 11:53		SY/MD	Ok,M
5	Q2011-02	IA1	VL042539.D	14 May 2025 12:40		SY/MD	Ok,M
6	Q2011-01	SV1	VL042540.D	14 May 2025 13:12		SY/MD	Ok,M
7	Q2037-02	IA1	VL042541.D	14 May 2025 14:15		SY/MD	Ok,M
8	Q2037-02DUP	IA1DUP	VL042542.D	14 May 2025 14:50		SY/MD	Ok,M
9	Q2037-01	SV1	VL042543.D	14 May 2025 15:22		SY/MD	Ok,M
10	Q2037-01DL	SV1DL	VL042544.D	14 May 2025 15:54	Not Required	SY/MD	Not Ok

M : Manual Integration