## ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

MATRIX: Water

METHOD: 8270-Modified/3510

1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	NA	NO	YES ✔
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			$\checkmark$
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			$\checkmark$
	The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		$\checkmark$	
7.	Surrogate Recoveries Meet Criteria		$\checkmark$	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for RW5-SP100-20250509 [Terphenyl-d14 - 265%], RW5-SP100-20250509DL [Terphenyl-d14 - 262%], RW5-SP201-20250509 [Terphenyl-d14 - 140%] and RW5-SP303-20250509 [Terphenyl-d14 - 148%], failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			$\checkmark$
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples .			

The Blank Spike Duplicate met requirements for all samples .

## ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## <u>GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY</u> (CONTINUED)

		NA	NO	YES			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			$\checkmark$			
	Comments:						
10.	Extraction Holding Time Met			$\checkmark$			
	If not met, list number of days exceeded for each sample:						
11.	Analysis Holding Time Met			$\checkmark$			
	If not met, list number of days exceeded for each sample:						
ADDITIONAL COMMENTS:							

The laboratory certifies that the all-electronic diskette deliverable exactly match the data

Summary forms (i.e. Form Is)."

Sample RW5-SP100-20250509 was diluted due to high concentration.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the % RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

**QA REVIEW** 

Date