

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051425\
 Data File : BN037006.D
 Acq On : 13 May 2025 22:29
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN051424

Quant Time: May 14 11:28:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.407	-1.7	98	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	105	0.00
4 S	2-Fluorophenol	0.400	0.423	-5.7	98	0.00
5 S	Phenol-d6	0.400	0.408	-2.0	97	0.00
6	bis(2-Chloroethyl)ether	0.400	0.383	4.3	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.379	5.3	105	0.00
9	Naphthalene	0.400	0.389	2.8	102	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.402	-0.5	104	0.00
12	2-Methylnaphthalene	0.400	0.396	1.0	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	104	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	104	0.00
16	Acenaphthylene	0.400	0.379	5.3	103	0.00
17	Acenaphthene	0.400	0.387	3.3	105	0.00
18	Fluorene	0.400	0.388	3.0	105	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.396	1.0	112	0.00
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	106	0.00
22	Hexachlorobenzene	0.400	0.411	-2.7	106	0.00
23	Atrazine	0.400	0.379	5.3	106	0.00
24	Pentachlorophenol	0.400	0.336	16.0	95	0.00
25	Phenanthrene	0.400	0.391	2.3	106	0.00
26	Anthracene	0.400	0.384	4.0	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	105	0.00
28	Fluoranthene	0.400	0.380	5.0	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	103	0.00
30	Pyrene	0.400	0.412	-3.0	105	0.00
31 S	Terphenyl-d14	0.400	0.415	-3.7	105	0.00
32	Benzo(a)anthracene	0.400	0.394	1.5	103	0.00
33	Chrysene	0.400	0.399	0.3	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.387	3.3	102	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.407	-1.7	99	0.00
37	Benzo(b)fluoranthene	0.400	0.391	2.3	99	0.00
38	Benzo(k)fluoranthene	0.400	0.410	-2.5	100	0.00
39 C	Benzo(a)pyrene	0.400	0.395	1.3	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.399	0.3	98	0.00
41	Benzo(g,h,i)perylene	0.400	0.409	-2.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN051424